

Draft Report on

***Fourth interlaboratory comparison
exercise for $\delta^2\text{H}$ and $\delta^{18}\text{O}$ analysis
of water samples (WICO2011)***

Prepared by

*M. Choudhry, P. Aggarwal, M. van Duren, L. Poltenstein, L.
Araguas and T. Kurttas*

*Isotope Hydrology Laboratory
Division of Physical and Chemical Sciences
Department of Nuclear Sciences and Applications
International Atomic Energy Agency*

Vienna, December 2011

CONTENTS

TABLE OF FIGURES	III
1. INTRODUCTION	1
2. SAMPLE PREPARATION AND REFERENCE VALUES	1
3. EVALUATION AND PRESENTATION OF THE REPORTED DATA	3
3.1. Statistical methods	3
3.2. Presentation of the results	4
3.3. Performance rating	17
3.4. Comparison of performance with previous exercises.....	17
4. COMPARISON OF RESULTS OF ANALYTICAL INSTRUMENTS	25
5. BASIC QUALITY ASSURANCE ISSUES	27
6. CONCLUSIONS	28
7. ACKNOWLEDGEMENTS	29
8. REFERENCES	29
APPENDIX I. LISTING OF RESULTS FOR ALL FOUR SAMPLES	30
APPENDIX II. LIST OF PARTICIPANTS	44
APPENDIX III. WICO2011 ANNOUNCEMENTS	57
APPENDIX IV. LABORATORY REPORTING SHEET	58

TABLE OF FIGURES

FIG. 1. S-shape plot for $\delta^{18}\text{O}$ values of sample IAEA-OH-13. Outliers as recognized during the statistical evaluation are marked by open symbols.....	7
FIG. 2. S-shape plot for $\delta^{18}\text{O}$ values of sample IAEA-OH-14. Outliers as recognized during the statistical evaluation are marked by open symbols.....	8
FIG. 3. S-shape plot for $\delta^{18}\text{O}$ values of sample IAEA-OH-15. Values outside the plot range are not shown. Outliers as recognized during the statistical evaluation are marked by open symbols.	9
FIG. 4. S-shape plot for $\delta^{18}\text{O}$ values on sample IAEA-OH-16. Outliers as recognized during the statistical evaluation are marked by open symbols.....	10
FIG. 5. S-shape plot for $\delta^2\text{H}$ values on sample IAEA-OH-13. Outliers as recognized during the statistical evaluation are marked by open symbols.....	11
FIG. 6. S-shape plot for $\delta^2\text{H}$ values on sample IAEA-OH-14. Outliers as recognized during the statistical evaluation are marked by open symbols.....	12
FIG. 7. S-shape plot for $\delta^2\text{H}$ values on sample IAEA-OH-15. Outliers as recognized during the statistical evaluation are marked by open symbols.....	13
FIG. 8. S-shape plot for $\delta^2\text{H}$ values on sample IAEA-OH-16. Outliers as recognized during the statistical evaluation are marked by open symbols.....	14
FIG. 9. $\delta^2\text{H}/\delta^{18}\text{O}$ plot of all reported laboratory mean values for the sample IAEA-OH-13. Not all values could be displayed using the selected scale.....	15
FIG. 10. $\delta^2\text{H}/\delta^{18}\text{O}$ plot of all reported laboratory mean values for the sample IAEA-OH-14. Not all values could be displayed using the selected scale.....	15
FIG. 11. $\delta^2\text{H}/\delta^{18}\text{O}$ plot of all reported laboratory mean values for the sample IAEA-OH-15. Not all values could be displayed using the selected scale.....	16
FIG. 12. $\delta^2\text{H}/\delta^{18}\text{O}$ plot of all reported laboratory mean values for the sample IAEA-OH-16. Not all values could be displayed using the selected scale.....	16
FIG. 13. Deviation of $\delta^{18}\text{O}$ results versus reference values for all four samples analyzed by each laboratory and sorted by increasing deviation.	18
FIG. 14. Deviation of $\delta^2\text{H}$ results versus reference values for all four samples analyzed by each laboratory and sorted by increasing deviation.	19
FIG. 15. Comparison between reported $\delta^{18}\text{O}$ uncertainty (lower part) and deviation from the reference value (upper part) for each laboratory.....	20
FIG. 16. Comparison between reported $\delta^2\text{H}$ uncertainty (lower part) and deviation from the reference value (upper part) for each laboratory.....	21
FIG. 17. Indicator for overall laboratory performance derived from both $\delta^2\text{H}$ and $\delta^{18}\text{O}$ data.	22
FIG. 18. Comparison of laboratory performance in WICO2011 with that of previous exercises for analysis of $\delta^{18}\text{O}$ in water samples.....	23
FIG. 19. Comparison of laboratory performance in WICO2011 with that of previous exercises for analysis of $\delta^2\text{H}$ in water samples.....	24
FIG. 20. Comparison of average values of $\delta^{18}\text{O}$ of four samples determined from the results of mass spectrometers and laser-based water isotope analyzers.....	26
FIG. 21. Comparison of average values of $\delta^2\text{H}$ of four samples determined from the results of mass spectrometers and laser-based water isotope analyzers.....	27

1. INTRODUCTION

The IAEA Isotope Hydrology Laboratory organized the fourth interlaboratory comparison exercise for laboratories engaged in routine analysis of hydrogen and oxygen stable isotope composition of water samples in 2011. Three similar exercises were carried out in 1995 [1], in 1999 [2] and in 2002 [3]. However, the tradition of IAEA water stable isotope inter-laboratory comparison is much older. Two interlaboratory comparison trials for isotope hydrology laboratories were carried out in the sixties and seventies, which revealed problems with use of the NBS-1 international standard; these data were used to calibrate the newly produced primary reference materials VSMOW and SLAP.

The WICO2011 exercise was announced in February 2011 on the internet, via the ISOGEOCHEM news group of Isotope geochemistry [4] and by email to all participants of the former intercomparisons.

Altogether 174 laboratories showed interest to participate in the exercise. Four water samples prepared and calibrated at the IAEA Isotope Hydrology Laboratory were labelled IAEA-OH-13 to IAEA-OH-16, which are referred to in this report as OH-13 to OH-16. By the end of the reporting deadline (the end of August 2011) altogether 137 laboratories from 53 countries had submitted 172 datasets back to the IAEA on the oxygen and hydrogen isotopic composition of these water samples.

The four water samples cover the range of $\delta^{18}\text{O}$ and $\delta^2\text{H}$ values typical for the majority of natural waters. The samples were bottled from 30 L stainless steel storage barrels into 30 mL securely-capped brown glass bottles, serially numbered at the time of filling. Each laboratory received a set of four samples with a corresponding code. This code (assigned randomly) forms the Identification (ID) code used throughout the exercise and in the tables and graphs of this report for each laboratory. The ID code is not related to the order of the list of participating laboratories. The identity of participating laboratories will not be revealed unless each laboratory explicitly agrees to do so.

2. SAMPLE PREPARATION AND REFERENCE VALUES

Waters collected from different areas were used to prepare four water samples for this intercomparison exercise. A brief description of these waters is provided below.

OH-13: Lake water from the shoreline of the lake Neusiedl, a shallow endorheic basin near Vienna, Austria. Fifty litres were filtered with a 5 μm filter and subsequently distilled using a Barnstead MP6A laboratory distiller.

OH-14: Commercially bottled water from the Moree wells in the great artesian basin in Australia. The water was not filtered or distilled.

OH-15: Commercially bottled water from the Libyan Al-Kufra region in the Nubian Sandstone Aquifer. The water was not filtered or distilled.

OH-16: Melted alpine snow sampled in the Salzburg region of the Austrian Alps at 2000 m.a.s.l. The water was not filtered or distilled.

The physical properties of these waters can be seen in Table 1.

TABLE 1. Physical properties of water samples used for intercomparison exercise

Name	pH	Conduct. $\mu\text{S}/\text{cm}$	Visual	Odour
OH-13	4.5	2	clear	no
OH-14	7.0	900	clear	no
OH-15	5.5	450	clear	no
OH-16	4.5	24	clear	no

After preparation, all four samples were stored in airtight 30 L stainless steel drums under slight argon overpressure. The containers are equipped with a special water extraction system allowing samples to be drawn without exposition to atmospheric air, thus avoiding any risk of evaporation/contamination. A detailed description of the system is available at <http://www-naweb.iaea.org/napc/ih/index.html>.

The water was distributed in 30 mL brown glass bottles, using an adapted Brand® bottle-top dispenser and a concentric double-pipe nozzle, flushed with argon gas. Bottles were capped securely and labelled with a sample name and filling sequence number.

The four water samples were analyzed for $\delta^{18}\text{O}$ and $\delta^2\text{H}$ at the Isotope Hydrology Laboratory by a dual inlet isotope ratio mass spectrometer (Finnigan Delta plus) equipped with a 48-port equilibration unit (for $\delta^{18}\text{O}$ and $\delta^2\text{H}$), and four laser based liquid water isotope analyzers (LWIA), three LGR and one Picarro. The mean isotopic values determined by mass spectrometer along with associated standard uncertainties u for a single measurement at the 1σ -level (Table 2) were sent to the participating laboratories initially as reference values. The mean values and uncertainties are derived from n accepted individual measurements as stated in the columns ‘number of analyses’ (number of rejected values in brackets).

TABLE 2. Isotope values for the four WICO2011 water samples as determined at the IAEA Isotope Hydrology Laboratory. Standard uncertainty u for a single analysis is given at the 1σ level. The number of accepted values to derive the mean value is reported (the number of the rejected values in brackets).

Sample	$\delta^{18}\text{O}$			$\delta^2\text{H}$		
	mean value [‰]	standard deviation 1σ [‰]	number of values	mean value [‰]	standard deviation 1σ [‰]	number of values
OH-13	-0.96	0.04	24	-2.29	0.94	23 (1)
OH-14	-5.59	0.05	23 (1)	-37.69	0.82	20 (4)
OH-15	-9.37	0.04	23 (1)	-78.01	0.77	20 (4)
OH-16	-15.41	0.04	24	-113.81	0.94	21 (3)

Later on, it was decided not to consider our results as reference values. Instead, the reference values were calculated from the results of the 12 laboratories whose performance was the best in the last intercomparison exercise, WICO2002, having cumulative deviations for $\delta^{18}\text{O}$ and $\delta^2\text{H}$ less than 0.2 ‰ and 2.0 ‰, respectively. For this purpose, their results in the 2011

exercise were processed according to the statistical methods provided in the subsequent section and the reference values are shown in Table 3.

TABLE 3. Isotope reference values for the four WICO2011 water samples as determined by the twelve laboratories which had best performance in the WICO2002 exercise with cumulative deviations for $\delta^{18}\text{O}$ and $\delta^2\text{H}$ of less than 0.2 ‰ and 2.0 ‰, respectively.

Sample	$\delta^{18}\text{O}$			$\delta^2\text{H}$		
	mean value [‰]	standard deviation 1σ [‰]	number of values	mean value [‰]	standard deviation 1σ [‰]	number of values
OH-13	-0.96	0.04	11	-2.84	0.60	12
OH-14	-5.60	0.05	12	-38.30	0.36	12
OH-15	-9.41	0.04	12	-78.26	0.38	11
OH-16	-15.43	0.04	12	-114.62	0.43	12

3. EVALUATION AND PRESENTATION OF THE REPORTED DATA

All the individual results of $\delta^{18}\text{O}$ and $\delta^2\text{H}$ submitted to the IAEA by the participating laboratories are listed in APPENDIX I. The list of participating laboratories is presented in APPENDIX II. The WICO2011 announcement and the Laboratory Reporting Sheet used in the exercise are presented in APPENDIX III and APPENDIX IV, respectively.

3.1. Statistical methods

Statistical analysis of the submitted results was undertaken with the major objective to assess the performance of individual laboratories (precision and accuracy) with respect to the reference $\delta^{18}\text{O}$ and $\delta^2\text{H}$ values for the analyzed four samples.

To reach this goal, a two-stage statistical treatment, adopted in previous IAEA interlaboratory exercises of a similar nature [1-5], was applied to the entire population of submitted results. All data evaluation and outlier determination were performed with commercial EXCEL spreadsheet programme.

The two-stage statistical treatment:

In stage I of the statistical treatment, obvious outliers were discarded based on the frequency distribution of values: after determination of the upper (H_U) and lower (H_L) quartiles, as well as the interquartile range ($H_U - H_L$), values exceeding $H_U + 3.0 \cdot (H_U - H_L)$ and $H_L - 3.0 \cdot (H_U - H_L)$ were discarded. The provisional mean compiled after stage I was further used for the stage II-outlier-rejection procedure.

In stage II of the evaluation process, the remaining results were assessed for each laboratory through examination of the difference between the reported isotope result x and the provisional mean m of stage I, divided by the standard uncertainty s quoted by the given laboratory. The results for which the ratio $|(x-m)/s|$ was larger than two were discarded, implying a significant deviation from the expected concordance of results. This procedure

identified those results which were seriously overestimating their measurements' precision.

In the final step, the weighted average was calculated by weighing the individual results by the reciprocal of the quoted variance:

$$X_w = \frac{\sum_{i=1}^n \frac{x_i}{s_i^2}}{\sum_{i=1}^n \frac{1}{s_i^2}} \quad (1)$$

The estimated standard uncertainty of the mean (ese) was calculated according to the following formula

$$ese(X_w) = \frac{\sigma_w}{\sqrt{\left(\sum_{i=1}^n \frac{1}{s_i^2}\right)}} \quad (2)$$

where:

$$\sigma_w^2 = \frac{\sum_{i=1}^n \frac{(x_i - X_w)^2}{s_i^2}}{n - 1} \quad (3)$$

where n stands for the final number of results accepted after the second step of the data evaluation procedure.

3.2. Presentation of the results

Before rejecting outliers using statistical methods, average of $\delta^{18}\text{O}$ and $\delta^2\text{H}$ values for the samples, along with associated standard deviations at 1σ -level and the total number of data were determined, are provided in Table 4.

TABLE 4. Average isotope values as determined by all participating laboratories for the four WICO2011 water samples.

Sample	$\delta^{18}\text{O}$			$\delta^2\text{H}$		
	mean value [‰]	standard deviation 1σ [‰]	number of values	mean value [‰]	standard deviation 1σ [‰]	number of values
OH-13	-1.01	0.27	166	-2.85	1.59	159
OH-14	-5.62	0.26	167	-38.47	1.52	160
OH-15	-9.41	0.25	167	-78.49	1.39	160
OH-16	-15.46	0.28	167	-114.99	1.76	160

Long term laboratory uncertainties were provided by only 45 laboratories; therefore stage II evaluation was undertaken using: (a) standard deviations associated with present measurements, (b) maximum of the measurement and long term standard deviations. In Table 5 the average values for the four samples are provided as evaluated from all data submitted by participating laboratories using measurement standard deviations in the evaluation scheme discussed in the previous section. The results obtained by using the measurement of maximum and long term standard deviations are depicted in Table 6. Both the tables provide the average values and associated standard deviations at 1 σ -level and the number of accepted laboratory values (the number of rejected laboratory values is in brackets). In the first case, some results with good accuracy were rejected in stage II due to overestimated precision, therefore Table 6 better represents the consensus values.

TABLE 5. Weighted average of accepted isotope values of all laboratories for the four WICO2011 water samples on the basis of reported measurement uncertainty. In the column ‘number of values’, the number of rejected values is stated in addition in brackets.

Sample	$\delta^{18}\text{O}$			$\delta^2\text{H}$		
	mean value [‰]	standard deviation 1 σ [‰]	number of values (rejected)	mean value [‰]	standard deviation 1 σ [‰]	number of values (rejected)
OH-13	-0.98	0.11	125(41)	-2.74	0.79	106(53)
OH-14	-5.59	0.10	128(39)	-38.40	0.82	117(43)
OH-15	-9.39	0.11	122(45)	-78.36	0.81	113(47)
OH-16	-15.45	0.11	123(44)	-114.94	0.70	109(51)

TABLE 6. Weighted average of accepted isotope values of all laboratories for the four WICO2011 water samples on the basis of maximum of stated and reported measurement uncertainty. In the column ‘number of values’, the number of rejected values is stated in addition in brackets.

Sample	$\delta^{18}\text{O}$			$\delta^2\text{H}$		
	mean value [‰]	standard deviation 1 σ [‰]	number of values (rejected)	mean value [‰]	standard deviation 1 σ [‰]	number of values (rejected)
OH-13	-0.97	0.12	140(26)	-2.73	0.91	123(36)
OH-14	-5.59	0.11	139(28)	-38.40	0.86	121(39)
OH-15	-9.39	0.13	132(35)	-78.31	0.87	126(34)
OH-16	-15.44	0.13	139(28)	-114.88	0.76	119(41)

The submitted laboratory mean values for $\delta^{18}\text{O}$ are listed in Table A1 in APPENDIX I, with columns listing first the assigned laboratory identification number (ID), and equipment used (mass spectrometer, LGR laser analyzer, Picarro laser analyzer, and in the following columns several laboratory parameters: the indication of a performed VSMOW/SLAP calibration

along the WICO2011 analyses, the number of laboratory water standards used for daily calibration and the amount of water and duration of the water/CO₂ equilibration. In the next columns the submitted mean values with associated standard uncertainties at a 1 σ -level are reported for samples OH-13 to OH-16. The grey shaded results were rejected in the statistical evaluation and were not used to calculate the weighted averages. In the following four figures (Figs 1 to 4) these data are displayed graphically as S-shape plots versus laboratory identification numbers sorted by increasing numerical values with uncertainties provided here at the 2 σ -level. Outliers having biases with respect to reference values greater than 2 times the sum of the standard deviations associated with the individual results (maximum) and the reference values at 1 σ -level are marked in the figures with open symbols. The reference value and its standard uncertainty at 2 σ -level are indicated in each figure as horizontal lines.

Similarly the submitted $\delta^2\text{H}$ results are listed in Table A2 in APPENDIX I, with columns listing assigned laboratory identification number (ID), equipment used (mass spectrometer, LGR laser analyzer, Picarro laser analyzer); and several laboratory parameters as the indication of a performed VSMOW/SLAP calibration, along with the WICO2011 analyses, the sample preparation method used, the number of laboratory water standards used for daily calibration and the amount of water used per measurement. In the following columns the submitted mean values with associated standard uncertainties at a 1 σ -level are reported for samples OH-13 to OH-16. In Figs 5 to 8, these data are displayed graphically as S-shape plots versus laboratory identification numbers sorted by increasing numerical values. Outliers as defined above are marked in the figures with open symbols.

A large spread of uncertainties is visible in the Figs 1 to 8; for a considerable number of laboratories the uncertainties are of such an order of magnitude, that measured data could not be used in a meaningful way in hydrological applications. This is recognized, but was not used as acceptance criterion in the evaluation, since the exercise was intentionally open to all laboratories measuring stable isotopes in water samples and was not limited to isotope hydrology laboratories.

A different display of the same data is provided by $\delta^2\text{H}/\delta^{18}\text{O}$ plots. These plots are quite useful to visualize the spread of hydrogen and oxygen data provided for any given sample. In Figs. 9 to 12, the reported laboratory mean data are plotted for each water sample. In the same plots, the respective average of the reported values calculated from all participating laboratories (except rejected values) is marked as a circle with associated uncertainty (1 σ -level). The squared symbol marks the reference value and its uncertainty (see Table 2). Several values could not be displayed in the plot at the chosen scales, but all the four plots cover a similar range of δ -values (with a range of 1.2 ‰ and 12 ‰ each). It is worthwhile to note that the average values for all laboratories are not much different from the reference values. For OH-13 sample, only $\delta^{18}\text{O}$ is slightly biased, while for OH-16 sample, both $\delta^{18}\text{O}$ and $\delta^2\text{H}$ are slightly biased towards more negative δ -values versus the reference values.




FIG. 1: S-shape plot for $\delta^{18}\text{O}$ values of sample IAEA-OH-13. Outliers as recognized during the statistical evaluation are marked by open symbols.




FIG. 2. S-shape plot for $\delta^{18}\text{O}$ values of sample IAEA-OH-14. Outliers as recognized during the statistical evaluation are marked by open symbols.




FIG. 3. S-shape plot for $\delta^{18}\text{O}$ values of sample IAEA-OH-15. Outliers as recognized during the statistical evaluation are marked by open symbols.




FIG. 4. S-shape plot for $\delta^{18}\text{O}$ values on sample IAEA-OH-16. Outliers as recognized during the statistical evaluation are marked by open symbols.




FIG. 5. S-shape plot for δ^2H values on sample IAEA-OH-13. Outliers as recognized during the statistical evaluation are marked by open symbols.




FIG. 6. S-shape plot for $\delta^2\text{H}$ values on sample IAEA-OH-14. Outliers as recognized during the statistical evaluation are marked by open symbols.




FIG. 7. S-shape plot for δ^2H values on sample IAEA-OH-15. Outliers as recognized during the statistical evaluation are marked by open symbols.




FIG. 8. S-shape plot for δ^2H values on sample IAEA-OH-16. Outliers as recognized during the statistical evaluation are marked by open symbols.




FIG. 9. $\delta^2\text{H}/\delta^{18}\text{O}$ plot of all reported laboratory mean values for the sample IAEA-OH-13. Not all values could be displayed using the selected scale.




FIG. 10. $\delta^2\text{H}/\delta^{18}\text{O}$ plot of all reported laboratory mean values for the sample IAEA-OH-14. Not all values could be displayed using the selected scale.




FIG. 11. $\delta^2\text{H}/\delta^{18}\text{O}$ plot of all reported laboratory mean values for the sample IAEA-OH-15. Not all values could be displayed using the selected scale.




FIG. 12. $\delta^2\text{H}/\delta^{18}\text{O}$ plot of all reported laboratory mean values for the sample IAEA-OH-16. Not all values could be displayed using the selected scale.

3.3. Performance rating

In order to obtain information on relative performance of the participating laboratories, the results were processed as they were in the last IAEA exercises [2, 3]. An obvious measure for the accuracy of a result is the deviation from the respective reference value. Since the measurement process for all four samples is the same and the precision of measurements should theoretically not be strongly dependent on this range of δ -values, the four sample results can be treated as providing independent tests for the accuracy of measurements in each laboratory. In Fig. 13 the cumulative deviation from the reference value for all four $\delta^{18}\text{O}$ results is displayed as a stack plot for each laboratory. A value equal to the reference value therefore has a value of zero on a $\Delta\delta^{18}\text{O}$ scale ($\Delta\delta^{18}\text{O}=\delta^{18}\text{O}_{\text{lab}}-\delta^{18}\text{O}_{\text{reference}}$). The laboratories are sorted in order of increasing absolute cumulative deviation. A similar approach was used for cumulative deviation from the reference value for all four $\delta^2\text{H}$ measurements and is shown in Fig. 14.

Deviations of the reported values from the reference values were also compared with the uncertainties to check the performance of a laboratory. As indicated above, most laboratories did not provide long term uncertainties, therefore reported measurement uncertainties were used in such cases. The cumulated deviation of $\delta^{18}\text{O}$ values from the reference values for all four values is plotted in Fig. 15. The thick black line indicates the laboratory uncertainty estimate (four times the standard uncertainty or sum of the measurement uncertainties for four samples). Columns much higher than the black line indicate an underestimation of uncertainty for those laboratories (deviations that are much larger than expected). On the other hand, columns being lower than the black line indicate an excessively conservative approach in uncertainty evaluation at a particular laboratory (real precision and accuracy better than assumed). A similar plot for $\delta^2\text{H}$ measurements is provided in Fig. 16.

For overall of laboratory performance assessment of both $\delta^{18}\text{O}$ and for $\delta^2\text{H}$, laboratory deviations as displayed in Fig. 15 and Fig. 16 were combined on the basis of their natural correlation by the mean meteoric water line [3]. The absolute value of all sample deviations for $\delta^2\text{H}$ was combined with similar data for $\delta^{18}\text{O}$, the latter data being multiplied by a factor of eight to weigh the deviations at a similar level. The resulting numerical value for each laboratory provides an indicator for laboratory deviation from reference values for the four distributed samples. A smaller value can be interpreted as better corresponding to the reference values. Those data are displayed in Fig. 17. Laboratories which only analyzed $\delta^{18}\text{O}$ or $\delta^2\text{H}$ are displayed separately at the right side of the plot. Such a performance indicator is somewhat arbitrary, as it combines different and independent measurements performed outside of the routine modus and provides only a snapshot of laboratory performance limited to the time period when the measurements were actually performed. No extrapolation is possible to the past and future based on the performance indicators presented here.

3.4. Comparison of performance with previous exercises

In order to compare the performance of the laboratories which participated in WICO2002 and/or WICO1999, cumulative deviations from the reference values for all four $\delta^{18}\text{O}$ and $\delta^2\text{H}$ results are plotted in the form of bar charts (Figs. 18 and 19). Most of the laboratories have improved their performance compared to the previous exercises.




FIG. 13. Deviation of $\delta^{18}\text{O}$ results versus reference values for all four samples analyzed by each laboratory and sorted by increasing deviation.




FIG. 14. Deviation of $\delta^2\text{H}$ results versus reference values for all four samples analyzed by each laboratory and sorted by increasing deviation.




FIG. 15. Comparison between reported $\delta^{18}\text{O}$ uncertainty and deviation from the reference value for laboratories. Columns much higher than the black line indicate an underestimation of uncertainty (deviations much greater than expected). On the other hand, columns much lower than the black line indicate an excessively conservative approach in uncertainty evaluation (real precision and accuracy better than assumed).




FIG. 16. Comparison between reported $\delta^2\text{H}$ uncertainty and deviation from the reference value for laboratories. Columns much higher than the black line indicate an underestimation of uncertainty (deviations much greater than expected). On the other hand, columns much lower than the black line indicate an excessively conservative approach in uncertainty evaluation (real precision and accuracy better than assumed).




FIG. 17. Indicator for overall laboratory performance derived from both $\delta^2\text{H}$ and $\delta^{18}\text{O}$ data.




FIG. 18. Comparison of laboratory performance in WICO2011 with that of previous exercises for analysis of $\delta^{18}\text{O}$ in water samples.




FIG. 19. Comparison of laboratory performance in WICO2011 with that of previous exercises for analysis of $\delta^2\text{H}$ in water samples.

4. COMPARISON OF RESULTS OF ANALYTICAL INSTRUMENTS

Results of three analytical instruments i.e. mass spectrometers, LGR laser-based water isotope analyzers and Picarro laser-based water isotope analyzers were compared. Average values of accepted results from WICO2011 samples determined by these instruments are provided in Tables 7 to Table 9. Graphic comparisons of average $\delta^{18}\text{O}$ and $\delta^2\text{H}$ values of the four samples determined from the results of these analyzers are shown in Fig. 20 and Fig. 21, respectively. It is noted that the average values for the three instruments are in very good agreement, being well within 1σ -uncertainties. The average values are so close that their maximum standard deviations at 1σ -level for $\delta^{18}\text{O}$ and $\delta^2\text{H}$ are 0.01‰ and 0.1‰, respectively.

TABLE 7. Weighted average of accepted isotope values analyzed by mass spectrometers of the four WICO2011 water samples on the basis of maximum stated and measurement uncertainty. In the column ‘number of laboratories’, the number of rejected laboratories is additionally stated in brackets.

Sample	$\delta^{18}\text{O}$			$\delta^2\text{H}$		
	Mean value [‰]	standard deviation 1σ [‰]	number of laboratories (rejected)	mean value [‰]	standard deviation 1σ [‰]	number of laboratories (rejected)
OH-13	-0.98	0.10	75(15)	-2.77	1.00	64(18)
OH-14	-5.60	0.10	72(18)	-38.37	0.94	67(15)
OH-15	-9.39	0.13	72(18)	-78.32	1.02	68(14)
OH-16	-15.44	0.13	71(19)	-114.83	0.81	66(16)

TABLE 8. Weighted average of accepted isotope values analyzed by LGR laser spectroscopic analyzers of the four WICO2011 water samples on the basis of maximum of stated and measurement uncertainty. In the column ‘number of laboratories’, the number of rejected laboratories is additionally stated in brackets.

Sample	$\delta^{18}\text{O}$			$\delta^2\text{H}$		
	mean value [‰]	standard deviation 1σ [‰]	number of laboratories (rejected)	mean value [‰]	standard deviation 1σ [‰]	number of laboratories (rejected)
OH-13	-0.97	0.18	34(7)	-2.68	0.75	32(9)
OH-14	-5.61	0.14	35(7)	-38.50	0.89	33(9)
OH-15	-9.41	0.13	33(9)	-78.46	0.71	30(12)
OH-16	-15.46	0.15	35(7)	-115.05	0.74	28(14)

TABLE 9. Weighted average of accepted isotope values analyzed by Picarro laser spectroscopic analyzers of the four WICO2011 water samples on the basis of maximum stated and measurement uncertainty. In the column ‘number of laboratories’, the number of rejected laboratories is additionally stated in brackets.

Sample	$\delta^{18}\text{O}$			$\delta^2\text{H}$		
	mean value [‰]	standard deviation 1σ [‰]	number of laboratories (rejected)	mean value [‰]	standard deviation 1σ [‰]	number of laboratories (rejected)
OH-13	-0.96	0.11	31(4)	-2.73	0.84	27(9)
OH-14	-5.58	0.09	32(3)	-38.40	0.47	21(15)
OH-15	-9.38	0.11	27(8)	-78.31	0.50	28(8)
OH-16	-15.45	0.09	33(2)	-114.89	0.62	25(11)




FIG. 20. Comparison of average $\delta^{18}\text{O}$ values of four samples determined from the results of mass spectrometers and laser-based water isotope analyzers.




FIG. 21. Comparison of average $\delta^2\text{H}$ values of four samples determined from the results of mass spectrometers and laser-based water isotope analyzers.

5. BASIC QUALITY ASSURANCE ISSUES AND RECOMMENDATIONS

The following are the main quality assurance issues:

Unacceptably high errors: Of all the participating laboratories, 35% for $\delta^{18}\text{O}$ and 39% for $\delta^2\text{H}$, have considerable problems in achieving the precision and accuracy targets required for most hydrological applications. Their cumulative biases for $\delta^{18}\text{O}$ and $\delta^2\text{H}$ are greater than 0.4‰ and 4‰, respectively.

Systematic errors: A general bias of reported results with respect to reference values exists towards both positive and negative values for the set of all participating laboratories. Inconsistencies in the results arise depending on the sample measurements carried out with aliquots of VSMOW and SLAP (for direct calibration) or using internal laboratory water standards calibrated earlier. Evaporated or contaminated standards result in systematic off-sets. Most of the laboratories have either positive or negative deviations for all four samples which cannot be attributed to random offsets in ^2H or ^{18}O measurements alone. One possible explanation for the shift towards more negative values could be slight evaporative enrichment of the used internal laboratory standards with time after calibration against VSMOW. Laboratory standards must be stored in proper containers with dispensing systems and slightly positive inert gas pressure to avoid evaporation and contamination by atmospheric moisture. Positive shifts could be due to evaporative enrichment of the primary standard (VSMOW), used multiple times for calibration of internal laboratory standards after the first opening of ampoules and improper storage.

Reporting of uncertainties: Stated long term and individual measurement uncertainties from some laboratories are too high or too low. The underestimation of uncertainty is especially critical. Other uncertainties stated are so huge that the results themselves become very uncertain and are of limited usefulness for most hydrological applications. Some laboratories quoted extremely low uncertainties averaging a few analytical measurements which are very close, especially regarding laser spectroscopic analysis. Overestimated uncertainties lead to the rejection of such results. Proper procedure should be used to reject outlying measurements for the calculation of average values and estimation of realistic uncertainties. Proper procedure should be used to reject outlying measurements for the calculation of average values and the estimation of realistic uncertainties.

Routine calibration with internal standards: It is obvious that for stable isotope measurements, daily routine calibration using only one internal laboratory standard cannot be used to correct for any scale normalization effect. Two internal laboratory standards would allow, for scale normalization. In the present exercise, seven laboratories used only one standard for calibration. The results of two to four samples of six of those laboratories are not within an acceptable range.

6. CONCLUSIONS

The number of participating laboratories in WICO2011 is twice the number of participants in WICO2002, mainly because of a revolution brought on by the expansion in the number of laser water isotope analyzers. It also shows increased interest in the application of environmental isotopes in hydrology and other disciplines.

No significant dependence of obtained δ -values on the analyzer type could be found. The results of laser water isotope analyzers and mass spectrometers are within the acceptable range of uncertainties.

Data from 52 sets out of 172 sets showed considerable deviation from the reference values for at least three sample values. On the other hand, a majority of laboratories have proven themselves capable of providing high quality data which can be used confidently for hydrological/ hydrogeological studies.

Precise calibration possible creates a bottleneck for those laboratories which are still subject to considerable offsets from the reference values. The basic requirement remains to suitably calibrate and store internal laboratory water standards.

The WICO2011 interlaboratory comparison has been helpful in detecting potential problem areas for laboratories. There is still a considerable number of laboratories with basic problems in accuracy, precision, and in stating justifiable measurement uncertainty. Further efforts are needed to improve this situation. Some improvements will be suggested to individual laboratories having large deviations in results and uncertainty problems.

Most of the laboratories which participated in WICO2002 and/or WICO1999 have improved their performance as compared to the previous exercises.

7. ACKNOWLEDGEMENTS

The efforts of all laboratories which participated in the 4th IAEA interlaboratory comparison exercise and contributed their results to the success of this test are highly appreciated.

8. REFERENCES

- [1] Araguás-Araguás, L., K. Rozanski (1995): IAEA interlaboratory comparison for deuterium and oxygen-18 analysis of precipitation samples. IAEA Report, Vienna, 33 pp.
- [2] Lippmann, J., M. Gröning, K. Rozanski (1999): 2nd interlaboratory comparison for deuterium and oxygen-18 analysis of water samples. IAEA Report, Vienna, 44 pp.
- [3] M. Gröning, S. Graeber, S. K. Bhattacharya, M. van Duren, L. Andreescu (2003): Third interlaboratory comparison exercise for δD and $\delta^{18}O$ analysis of water samples (WICO2002-Part 1). IAEA Report, Vienna, 43 pp.
- [4] ISOGEOCHEM newsgroup: <http://geology.uvm.edu/geowww/isogeochem.html>
- [5] Rozanski, K., W. Stichler, R. Gonfiantini, E.M. Scott, R.P. Beukens, B. Kromer, J. Van der Plicht, (1992): The IAEA ¹⁴C intercomparison exercise. Radiocarbon 34 (3), 506-519.

APPENDIX I. LISTING OF RESULTS FOR ALL FOUR SAMPLES

Table A1: Submitted $\delta^{18}\text{O}$ results from all participating laboratories for samples OH-13 to OH-16 with measurement uncertainties at the 1σ -level. Gray-shaded values are regarded as outliers according to the statistical criteria applied. Column 'ID' gives the assigned laboratory identification number; column 'Equip' indicates equipment (mass spectrometer, LGR laser analyzer, Picarro laser analyzer) used; column 'CA' states in Yes/No modus, whether a calibration with VSMOW was performed along with WICO2011 measurements; 'NS' provides the number of laboratory water standards used for daily calibration; 'AW' shows the amount of water used for a single measurement; 'ET' provides the used equilibration time for the $\text{H}_2\text{O}/\text{CO}_2$ exchange reaction as used where applicable; 'Stated Unc.' indicates the long term laboratory uncertainties. Other parameters provided by the laboratories are not listed in this report. The abbreviations used in the columns: NA: not applicable, NM: not mentioned, IRMS: isotope ratio mass spectrometer, LGR: Los Gatos Research laser analyzer, Pic: Picarro laser analyzer.

Lab ID	Instrument	CA	NS	AW (μL)	ET (hr)	IAEA- OH-13 $\delta^{18}\text{O}$ 1 σ [‰]		IAEA- OH-14 $\delta^{18}\text{O}$ 1 σ [‰]		IAEA- OH-15 $\delta^{18}\text{O}$ 1 σ [‰]		IAEA- OH-16 $\delta^{18}\text{O}$ 1 σ [‰]		Stated Unc.	Remarks
Reference values						-0.96	0.04	-5.60	0.05	-9.41	0.04	-15.43	0.04		
1A	LGR	N	5	0.70	NA	-0.75	0.36	-4.96	0.18	-8.70	0.15	-15.16	0.14		
1B	IRMS	N	1	1000	25	-0.92	0.05	-5.89	0.11	-9.27	0.14	-15.72	0.14	0.20	
3	IRMS	NM	2	0.10	NM	-1.02	0.04	-5.57	0.05	-9.28	0.01	-15.33	0.05	0.15	
4	LGR	NM	1	NM	NA	-1.58	0.21	-5.46	0.17	-9.08	0.18	-13.57	0.12		
5	IRMS	Y	1	1000	NM	-1.01	0.07	-5.56	0.09	-9.31	0.05	-15.25	0.08		
6A	IRMS	Y	3	5000	10	-0.98	0.07	-5.58	0.03	-9.39	0.03	-15.42	0.02		
6B	PIC	Y	NM	1.75	NA	-0.89	0.11	-5.58	0.08	-9.27	0.17	-15.55	0.07		
7	IRMS	Y	2	3000	4	-1.13	0.09	-5.70	0.10	-9.53	0.10	-15.58	0.12		
8	IRMS	N	3	400	4	-1.09	0.29	-5.31	0.23	-9.26	0.25	-15.27	0.29	0.15	
9	IRMS	Y	3	0.70	NM	-0.90	0.29	-5.65	0.29	-9.65	0.41	-15.32	0.29		
11	LGR	NM	6	1000	NA	-0.88	0.06	-5.56	0.05	-9.37	0.05	-15.50	0.04	0.15	
13A	IRMS	Y	3	0.50	NM	-1.16	0.40	-5.53	0.34	-9.33	0.31	-15.28	0.46		
13B	IRMS	Y	3	0.50	18	-1.04	0.15	-5.61	0.13	-9.36	0.13	-15.37	0.25		

Lab ID	Instrument	CA	NS	AW (μL)	ET (hr)	IAEA- OH-13		IAEA- OH-14		IAEA- OH-15		IAEA- OH-16		Stated Unc.	Remarks
						$\delta^{18}\text{O}$	1 σ [‰]	$\delta^{18}\text{O}$	1 σ [‰]	$\delta^{18}\text{O}$	1 σ [‰]	$\delta^{18}\text{O}$	1 σ [‰]		
Reference values						-0.96	0.04	-5.60	0.05	-9.41	0.04	-15.43	0.04		
15A	IRMS	NM	4	5000	5	-0.93	0.02	-5.56	0.05	-9.36	0.04	-15.34	0.02	0.18	
15B	PIC	NM	NM	NM	NA	-0.96	0.05	-5.52	0.04	-9.32	0.06	-15.41	0.10		
19	LGR	N	3	0.97	NA	-0.98	0.06	-5.73	0.10	-9.51	0.08	-15.59	0.14		
22	LGR	N	2	NM	NA	-1.03	0.21	-5.69	0.19	-9.50	0.09	-15.66	0.13		
24	IRMS	Y	3	1.00	NA	-1.46	0.20	-6.06	0.21	-9.79	0.21	-15.70	0.22		
25A	PIC	Y	3	2.00	NA	-0.92	0.05	-5.63	0.05	-9.41	0.03	-15.50	0.04		
25B	IRMS	Y	3	NM	18	-0.97	0.04	-5.55	0.03	-9.38	0.02	-15.48	0.02		
26	IRMS	NM	3	2000	18	-0.88	0.07	-5.29	0.07	-9.27	0.05	-15.28	0.07		
27A	IRMS	N	3	200	18	-1.10	0.08	-5.52	0.05	-9.44	0.04	-15.54	0.14		
28	IRMS	Y	2	NM	NA	-0.83	0.20	-5.53	0.25	-9.35	0.26	-15.50	0.19		
29	IRMS	NM	NM	1.00	NA	-1.08	0.14	-5.84	0.12	-9.59	0.13	-15.65	0.15		
30	LGR	NM	3	NM	NA	-0.91	0.06	-5.62	0.11	-9.46	0.07	-15.46	0.15		
31	PIC	Y	NM	NM	NA	-1.07	0.11	-5.71	0.11	-9.52	0.16	-15.55	0.11		
34	IRMS	Y	3	300	20	-0.98	0.10	-5.51	0.11	-9.41	0.11	-15.42	0.11		
35	PIC	NM	2	10	NA	-1.25	0.33	-5.82	0.23	-9.55	0.28	-15.65	0.28	0.39	
36	PIC	Y	3	1.00	NA	-1.03	0.02	-5.60	0.02	-9.38	0.04	-15.38	0.02	0.20	
37	IRMS	Y	2	0.15	NA	-1.30	0.15	-5.80	0.21	-10.00	0.13	-16.00	0.14	0.33	
38A	PIC	N	3	1.80	NA	-0.98	0.08	-5.65	0.15	-9.55	0.09	-15.55	0.09		
38B	IRMS	N	3	0.50	18	-1.11	0.03	-5.72	0.09	-9.55	0.05	-15.64	0.02		
39	IRMS	Y	3	200	34	-1.11	0.07	-5.61	0.08	-9.58	0.09	-15.76	0.06		
40	LGR	Y	2	1.00	NA	-0.68	0.03	-5.57	0.19	-9.42	0.20	-15.43	0.19		
41	PIC	Y	2	1.77	NA	-0.94	0.02	-5.59	0.14	-9.40	0.04	-15.49	0.06	0.06	
43	PIC	N	2	1.80	NA	-1.09	0.10	-5.65	0.07	-9.41	0.04	-15.42	0.06	0.08	
44	IRMS	Y	3	200	7	-1.176	0.05	-5.67	0.02	-9.47	0.03	-15.50	0.05		
45	IRMS	N	7	2500	8	-0.989	0.02	-5.64	0.01	-9.43	0.01	-15.47	0.02		

Lab ID	Instrument	CA	NS	AW (μ L)	ET (hr)	IAEA- OH-13		IAEA- OH-14		IAEA- OH-15		IAEA- OH-16		Stated Unc.	Remarks
						$\delta^{18}\text{O}$	1 σ [‰]	$\delta^{18}\text{O}$	1 σ [‰]	$\delta^{18}\text{O}$	1 σ [‰]	$\delta^{18}\text{O}$	1 σ [‰]		
Reference values						-0.96	0.04	-5.60	0.05	-9.41	0.04	-15.43	0.04		
47A	IRMS	NM	4	4000	4	-0.970	0.04	-5.59	0.04	-9.42	0.05	-15.43	0.03	0.05	
48	IRMS	N	6	600	36	-0.94	0.05	-5.64	0.06	-9.45	0.06	-15.47	0.07		
50A	IRMS	Y	3	4000	4	-0.88	0.08	-5.53	0.06	-9.35	0.08	-15.40	0.06	0.06	
50B	PIC	NM	NM	1.70	NA	-0.86	0.03	-5.53	0.03	-9.32	0.04	-15.32	0.07	0.08	
51	LGR	N	3	1	NA	-0.86	0.06	-5.51	0.10	-9.41	0.13	-15.49	0.05		
52	IRMS	N	3	200	24	-1.03	0.02	-5.62	0.01	-9.42	0.02	-15.60	0.02	0.20	
53A	IRMS	NM	3	500	20	-0.82	0.05	-5.42	0.06	-9.22	0.08	-15.29	0.07	0.13	
53B	PIC	NM	3	10	NA	-0.77	0.08	-5.40	0.14	-9.17	0.11	-15.28	0.06	0.19	
56	PIC	NM	3	NM	NA	-0.87	0.07	-5.48	0.07	-9.32	0.06	-15.34	0.09	0.10	
57	IRMS	NM	2	500	20	-1.11	0.06	-5.63	0.04	-9.44	0.08	-15.35	0.09		
58	IRMS	Y	3	NM	NA	-0.83	0.05	-5.34	0.16	-9.13	0.16	-15.17	0.00	0.30	
59	IRMS	Y	3	200	12	-1.08	0.10	-5.65	0.09	-9.46	0.09	-15.51	0.06	0.10	
60A	IRMS	NM	2	1.00	NA	-0.89	0.00	-5.51	0.01	-9.30	0.01	-15.41	0.03		
60B	PIC	NM	2	NM	NA	-0.96	0.02	-5.68	0.12	-9.43	0.01	-15.57	0.10		
61	IRMS	Y	2	200	4	-0.74	0.20	-5.19	0.14	-8.85	0.08	-14.78	0.25		
62	IRMS	NM	1	5000	8	-0.95	0.17	-5.2	0.08	-8.94	0.22	-15.41	0.28		
63A	IRMS	N	4	500	NM	-0.93	0.06	-5.53	0.06	-9.39	0.06	-15.43	0.06		
63B	PIC	N	4	1.85	NA	-0.98	0.10	-5.61	0.09	-9.48	0.09	-15.40	0.10		
64A	IRMS	NM	2	500	24	-0.92	0.05	-5.56	0.26	-9.20	0.04	-15.20	0.06		
64B	IRMS	NM	2	0.70	NA	-0.980	0.16	-5.30	0.14	-9.04	0.18	-15.30	0.13		
66	IRMS	N	3	2000	2	-0.90	0.03	-5.52	0.06	-9.33	0.03	-15.39	0.02		
67	LGR	NM	3	1.00	NA	-0.85	0.10	-5.50	0.08	-9.41	0.08	-15.61	0.07		
69	LGR	NM	3	0.75	NA	-1.47	0.30	-5.81	0.23	-9.65	0.27	-15.56	0.21		
70	IRMS	Y	3	1.00	NA	-0.94	0.10	-5.59	0.10	-9.42	0.10	-15.54	0.10	0.10	
71	LGR	NM	NM	NM	NA	-1.02	0.13	-5.65	0.06	-9.57	0.09	-15.50	0.12		

Lab ID	Instrument	CA	NS	AW (μL)	ET (hr)	IAEA- OH-13 $\delta^{18}\text{O}$ $1\sigma[\text{‰}]$		IAEA- OH-14 $\delta^{18}\text{O}$ $1\sigma[\text{‰}]$		IAEA- OH-15 $\delta^{18}\text{O}$ $1\sigma[\text{‰}]$		IAEA- OH-16 $\delta^{18}\text{O}$ $1\sigma[\text{‰}]$		Stated Unc.	Remarks
Reference values						-0.96	0.04	-5.60	0.05	-9.41	0.04	-15.43	0.04		
72	IRMS	Y	3	200	4	-1.96	0.06	-6.61	0.06	-10.44	0.05	-16.40	0.06		
74A	IRMS	NM	3	NM	6	-1.00	0.03	-5.51	0.03	-9.34	0.03	-15.43	0.03	0.10	
75A	IRMS	N	3	400	24	-0.95	0.01	-5.62		-9.41		-15.50			
75B	LGR	N	3	0.92	NA	-0.91	0.09	-5.63	0.03	-9.48	0.03	-15.71	0.15		
77		NM	NM	NM	NM	-1.13	0.10	-5.90	0.10	-9.50	0.10	-15.47	0.10		
78	LGR	NM	3	NM	NA	-0.93	0.23	-5.66	0.25	-9.35	0.16	-15.51	0.15		
79	IRMS	N	2	500	18	-1.04	0.06	-5.66	0.08	-9.43	0.08	-15.44	0.05		
80A	IRMS	Y	NM	5000	12	-0.94	0.05	-5.57	0.05	-9.39	0.05	-15.46	0.05		
80B	PIC	Y	NM	1.80	NA	-1.00	0.10	-5.55	0.10	-9.39	0.10	-15.46	0.10		
81	PIC	Y	2	NM	NA	-0.89	0.14	-5.59	0.12	-9.37	0.18	-15.39	0.13		
82	LGR	N	5	0.75	NA	-0.70	0.19	-5.29	0.04	-9.15	0.15	-15.28	0.08	0.17	
84	IRMS	N	2	500	18	-1.00	0.13	-5.54	0.04	-9.47	0.04	-15.55	0.05	0.04	
86A	LGR	Y	2	NM	NA	-1.11	0.30	-5.65	0.30	-9.47	0.30	-15.51	0.30		
86B	IRMS	Y	2	500	55	-1.04	0.20	-5.64	0.20	-9.43	0.20	-15.49	0.20		
87	IRMS	Y	2	NM	24	-0.80	0.09	-5.90	0.05	-9.29	0.05	-15.34	0.06		
88A	IRMS	N	5	400	6	-1.0726	0.07	-5.58	0.07	-9.41	0.12	-15.44	0.06	0.10	
88B	IRMS	N	5	200	6	1.04	0.02	-5.70	0.02	-9.43	0.01	-15.56	0.06		
89	IRMS	N	3	200	7	-1.01	0.05	-5.60	0.02	-9.40	0.01	-15.46	0.01	0.04	
91	LGR	N	3	NM	NA	-0.84	0.12	-5.47	0.08	-9.24	0.07	-15.29	0.05		
93	IRMS	Y	3	1000	9	-1.00	0.02	-5.63	0.02	-9.41	0.02	-15.44	0.01		
95	IRMS	NM	1	NM	NA	-1.28	0.18	-5.57	0.27	-9.25	0.15	-15.11	0.26	0.16	
96	IRMS	N	2	1500	6	-0.94	0.03	-5.60	0.03	-9.38	0.04	-15.44	0.04	0.06	
98	PIC	NM	4	2.00	NA	-0.93	0.06	-5.57	0.02	-9.35	0.02	-15.44	0.07		
99	IRMS	N	2	2000	10	-1.00	0.03	-5.63	0.05	-9.46	0.15	-15.47	0.06		
100	IRMS	N	5	3000	21	-0.92	0.09	-5.55	0.07	-9.29	0.09	-15.40	0.03	0.10	

Lab ID	Instrument	CA	NS	AW (μL)	ET (hr)	IAEA- OH-13		IAEA- OH-14		IAEA- OH-15		IAEA- OH-16		Stated Unc.	Remarks
						$\delta^{18}\text{O}$	1 σ [‰]	$\delta^{18}\text{O}$	1 σ [‰]	$\delta^{18}\text{O}$	1 σ [‰]	$\delta^{18}\text{O}$	1 σ [‰]		
Reference values						-0.96	0.04	-5.60	0.05	-9.41	0.04	-15.43	0.04		
101	IRMS	Y	2	200	NM	-1.03	0.06	-5.64	0.04	-9.46	0.04	-15.49	0.06	0.10	
102	LGR	N	3	NM	NA	-1.59	0.09	-5.74	0.10	-9.29	0.05	-15.40	0.04	0.12	
103	IRMS	NM	5	1.00	24	-0.94	0.13	-5.62	0.09	-9.37	0.09	-15.59	0.11		
104	IRMS	Y	3	4000	9	-0.95	0.03	-5.59	0.01	-9.37	0.01	-15.42	0.02	0.02	
105	LGR	Y	3	NM	NA	-1.07	0.12	-5.74	0.08	-9.47	0.05	-15.54	0.06		
106	LGR	Y	2	NM	NA	-0.66	0.35	-5.89	0.45	-9.50	0.41	-15.31	0.19		
108	IRMS	NM	3	NM	4	-0.98	0.04	-5.58	0.05	-9.40	0.03	-15.42	0.02		
109	LGR	NM	3	NM	NA	-0.95	0.18	-5.63	0.15	-9.42	0.15	-15.41	0.13		
111	IRMS	Y	2	200	24	-0.99	0.08	-5.67	0.06	-9.51	0.03	-15.55	0.08	0.15	
112A	IRMS	N	3	3000	6	-1.05	0.06	-5.61	0.03	-9.37	0.03	-15.39	0.06		
112B	LGR	N	3	NM	NA	-1.06	0.17	-5.66	0.17	-9.40	0.14	-15.43	0.14		
113A	IRMS	Y	4	NM	18	-0.87	0.11	-5.28	0.18	-9.26	0.14	-15.36	0.20		
113B	PIC	Y	4	NM	NA	-0.96	0.02	-5.60	0.05	-9.32	0.10	-15.43	0.10		
114	LGR	Y	3	NM	NA	-1.24	0.13	-6.17	0.14	-9.91	0.11	-15.76	0.09		
116	IRMS	Y	3	4000	4	-0.98	0.14	-5.63	0.15	-9.43	0.15	-15.49	0.14	0.10	
118A	IRMS	Y	2	2000	12	-0.99	0.11	-5.63	0.14	-9.40	0.11	-15.62	0.12		
118B	PIC	NM	2	NM	NA	-1.08	0.16	-5.69	0.01	-9.52	0.03	-15.51	0.06		
119A	LGR	N	2	NM	NA	-1.02	0.28	-5.74	0.31	-9.29	0.37	-15.12	0.32		
119B	LGR	NM	NM	NM	NA			-5.06	0.42	-8.85	0.31	-15.34	0.26		
119C	LGR	NM	NM	NM	NA	-0.81	0.21	-5.47	0.12	-9.26	0.13	-15.33	0.25		
120	IRMS	NM	2	NM	12	-0.99	0.02	-5.62	0.02	-9.37	0.03	-15.43	0.01		
122	LGR	N	3	0.75	NA	-0.80	0.10	-5.60	0.30	-9.30	0.20	-15.40	0.10		
123	LGR	Y	3	1.20	NA	-0.92	0.22	-5.54	0.18	-9.48	0.19	-15.55	0.19		
125	LGR	N	3	0.75	NA	-0.78	0.06	-5.33	0.04	-9.21	0.05	-15.36	0.08	0.08	
126	IRMS	N	3	200	5	-0.92	0.12	5.37	0.06	-9.38	0.18	-15.39	0.10		

Lab ID	Instrument	CA	NS	AW (μ L)	ET (hr)	IAEA- OH-13		IAEA- OH-14		IAEA- OH-15		IAEA- OH-16		Stated Unc.	Remarks
						$\delta^{18}\text{O}$	1 σ [‰]	$\delta^{18}\text{O}$	1 σ [‰]	$\delta^{18}\text{O}$	1 σ [‰]	$\delta^{18}\text{O}$	1 σ [‰]		
Reference values						-0.96	0.04	-5.60	0.05	-9.41	0.04	-15.43	0.04		
127	IRMS	Y	2	3000	5	-0.78	0.10	-5.56	0.07	-9.33	0.05	-15.30	0.05		
129	LGR	NM	3	NM	NA	-0.88	0.10	-5.50	0.10	-9.40	0.08	-15.60	0.13	0.20	
130	LGR	NM	3	NM	NA	-1.23	0.09	-5.74	0.08	-9.43	0.15	-15.34	0.07	0.30	
132	LGR	N	3	1.20	NA	-0.95	0.25	-5.58	0.28	-9.34	0.33	-15.38	0.32		
133A	IRMS	N	3	0.10	NM	-1.03	0.09	-5.67	0.06	-9.58	0.07	-15.70	0.04		
133B	IRMS	NM	NM	NM	NA	0.13	0.09	-5.85	0.12	-9.86	0.06	-16.09	0.03		
135A	PIC	Y	3	2.00	NA	-0.95	0.01	-5.59	0.01	-9.37	0.01	-15.44	0.04		
135B	PIC	Y	3	2.00	NA	-0.84	0.041	-5.59	0.07	-9.35	0.01	-15.39	0.04		
136	IRMS	N	2	2000	11	-0.95	0.08	-5.61	0.06	-9.41	0.08	-15.47	0.08		
137	LGR	N	2	1.20	NA	-0.89	0.15	-5.33	0.20	-9.01	0.17	-15.06	0.23		
138	PIC	NM	NM	NM	NA	-0.88	0.11	-5.52	0.13	-9.38	0.09	-15.40	0.07		
139	IRMS	NM	2	NM	2	-1.03	0.01	-5.62	0.07	-9.38	0.04	-15.33	0.02		
140A	IRMS	NM	2	500	24	-0.91	0.14	-5.59	0.12	-9.40	0.37	-15.56	0.12	0.20	
141	IRMS	N	2	200	24	-0.96	0.04	-5.58	0.05	-9.39	0.08	-15.44	0.06	0.10	
142	LGR	N	2	NM	NA	-1.18	0.15	-5.68	0.13	-9.45	0.07	-15.47	0.33		
143	LGR	Y	3	NM	NA	-1.05	0.04	-6.53	0.05	-9.48	0.04	-16.03	0.06		
144	IRMS	N	3	1000	9	-0.96	0.02	-5.62	0.02	-9.41	0.01	-15.51	0.02		
145	LGR	NM	7	1.00	NA	-1.22	0.15	-5.76	0.11	-9.60	0.13	-15.63	0.13		
146A	IRMS	N	1	3000	24	-1.10	0.05	-5.76	0.07	-9.56	0.05	-15.61	0.02	0.10	
146B	PIC	N	3	1.00	NA	-1.17	0.02	-5.62	0.10	-9.60	0.07	-15.64	0.06	0.10	
146C	PIC	N	3	1.00	NA	-1.13	0.04	-5.85	0.11	-9.67	0.15	-15.71	0.05	0.10	
147A	IRMS	N	3	5000	10	-0.96	0.03	-5.60	0.02	-9.40	0.02	-15.45	0.03		
147B	PIC	N	3	2.00	NA	-1.01	0.06	-5.56	0.05	-9.40	0.12	-15.46	0.12		
148	IRMS	N	2	5000	10	-1.00	0.03	-5.62	0.08	-9.42	0.12	-15.50	0.04	0.10	
149A	PIC	Y	3	1.80	NA	-0.95	0.03	-5.57	0.02	-9.36	0.02	-15.41	0.03		
149B	IRMS	Y	3	3000	4	-0.97	0.02	-5.59	0.02	-9.36	0.03	-15.43	0.02		

Lab ID	Instrument	CA	NS	AW (μ L)	ET (hr)	IAEA- OH-13 $\delta^{18}\text{O}$ 1 σ [‰]		IAEA- OH-14 $\delta^{18}\text{O}$ 1 σ [‰]		IAEA- OH-15 $\delta^{18}\text{O}$ 1 σ [‰]		IAEA- OH-16 $\delta^{18}\text{O}$ 1 σ [‰]		Stated Unc.	Remarks			
Reference values						-0.96	0.04	-5.60	0.05	-9.41	0.04	-15.43	0.04					
152	IRMS	NM	3	1000	8	-0.91	0.06	-5.61	0.06	-9.45	0.10	-15.47	0.07	0.01				
153	IRMS	Y	3	400	7	-0.98	0.07	-5.52	0.08	-9.39	0.06	-15.40	0.04					
154	IRMS	NM	3	500	20	-1.00	0.10	-5.48	0.10	-9.13	0.15	-15.15	0.15					
155A	IRMS	Y	3	4000	NM	-0.96	0.04	-5.59	0.05	-9.37	0.04	-15.41	0.04					
155B	PIC	Y	3	1.00	NA	-1.05	0.13	-5.54	0.07	-9.34	0.06	-15.40	0.07					
155C	LGR	Y	3	0.75	NA	-1.05	0.16	-5.53	0.13	-9.37	0.12	-15.41	0.14					
155D	LGR	Y	3	0.75	NA	-0.78	0.18	-5.37	0.22	-9.14	0.21	-15.27	0.22					
155E	LGR	Y	3	0.75	NA	-0.74	0.29	-5.40	0.31	-9.20	0.21	-15.19	0.24					
156	IRMS	NM	2	1.00	NM	-1.28	0.11	-5.88	0.11	-9.77	0.05	-15.91	0.13			0.10		
157	IRMS	Y	1	NM	15	-1.00	0.16	-5.77	0.09	-9.65	0.09	-15.71	0.05					
159	IRMS	Y	2	NM	18	-0.88	0.10	-5.24	0.01	-9.26	0.06	-15.24	0.10					
160	PIC	Y	10	1.00	NA	-1.00	0.04	-5.63	0.05	-9.45	0.08	-15.46	0.04					
161	LGR	Y	3	NM	NA	-0.88	0.01	-5.51	0.02	-9.34	0.03	-15.41	0.01			0.04	0.20	
162	IRMS	N	4	200	5	-1.00	0.17	-5.50	0.21	-9.40	0.13	-15.30	0.10					
165	PIC	NM	3	NM	NA	-1.00	0.10	-5.59	0.03	-9.40	0.05	-15.46	0.05					
168	IRMS	Y	4	5000	4	-1.05	0.02	-5.59	0.05	-9.38	0.04	-15.40	0.05					
170	PIC	N	3	1.70	NA	-0.86	0.09	-5.48	0.13	-9.32	0.03	-15.36	0.10					
172	PIC	Y	3	1.70	NA	-1.25	0.14	-5.83	0.15	-9.63	0.15	-15.63	0.09					
173	IRMS	N	3	200	10	-0.61	0.06	-5.26	0.17	-9.15	0.14	-15.30	0.15					
175	LGR	NM	3	1.20	NA	-3.34	0.65	-7.64	0.57	-11.27	0.61	-17.21	0.70					
177	LGR	NM	3	NM	NA	-1.08	0.22	-5.68	0.20	-9.41	0.22	-15.57	0.19					
178	PIC	NM	3	NM	NA	-0.86	0.06	-5.51	0.05	-9.34	0.06	-15.33	0.08					
180	PIC	NM	NM	1.10	NA	-2.01	0.09	-6.69	0.04	-10.47	0.05	-16.52	0.03	0.30	Manual Injection			
181	LGR	NM	3	NM	NA	-0.70	0.09	-5.69	0.25	-9.27	0.20	-15.25	0.12					
182A	IRMS	Y	2	5000	5	-1.21	0.04	-5.64	0.04	-9.53	0.31	-15.31	0.01			0.20		

Lab ID	Instrument	CA	NS	AW (μL)	ET (hr)	IAEA- OH-13		IAEA- OH-14		IAEA- OH-15		IAEA- OH-16		Stated Unc.	Remarks
						$\delta^{18}\text{O}$	1 σ [‰]	$\delta^{18}\text{O}$	1 σ [‰]	$\delta^{18}\text{O}$	1 σ [‰]	$\delta^{18}\text{O}$	1 σ [‰]		
Reference values						-0.96	0.04	-5.60	0.05	-9.41	0.04	-15.43	0.04		
182B	LGR	Y	2	0.75	NA	-1.01	0.03	-5.65	0.09	-9.41	0.07	-15.54	0.11	0.20	
183	PIC	Y	NM	1.80	NA	-0.91	0.06	-5.64	0.07	-9.43	0.05	-15.47	0.08		
185	IRMS	NM	2	500	16	-0.99	0.07	-5.26	0.09	-9.16	0.04	-15.11	0.12		

Table A2: Submitted $\delta^2\text{H}$ results from all participating laboratories for samples OH-13 to OH-16 with measurement uncertainties at the 1 σ -level. Gray-shaded values are regarded as outliers according to the statistical criteria applied. Column ‘ID’ gives the assigned laboratory identification number; column ‘Equip’ indicates equipment (mass spectrometer, LGR laser analyzer, Picarro laser analyzer) used; column ‘CA’ states in Yes/No modus, whether a calibration with VSMOW was performed along with WICO2011 measurements; ‘Method’ provides the indication of the sample preparation: e.g. reduction methods (Zn for Zinc, Cr for Chromium), pyrolysis (Pyro), equilibration of H₂ gas with Pt-catalyst; the column ‘NS’ provides the number of laboratory water standards used for daily calibration; ‘AW’ shows the amount of water used for a single measurement; ET’ provides the used equilibration time for the H₂O/H₂ exchange reaction where applicable; ‘Stated Unc’ indicates the long term uncertainties. Other parameters provided by the laboratories are not listed in this report. The abbreviations used in the columns are: NA: not applicable, NM: not mentioned, IRMS: isotope ratio mass spectrometer, LGR: Los Gatos Research laser analyzer, Pic: Picarro laser analyzer.

Lab ID	Equip.	CA	Method	NS	AW (μL)	ET (hr)	IAEA-OH-13		IAEA-OH-14		IAEA-OH-15		IAEA-OH-16		Stated Unc.	
							$\delta^2\text{H}$	1 σ [‰]	$\delta^2\text{H}$	1 σ [‰]	$\delta^2\text{H}$	1 σ [‰]	$\delta^2\text{H}$	1 σ [‰]		
Reference values							-2.84	0.60	-38.30	0.36	-78.26	0.38	-114.62	0.43		
1a	LGR	N	NA	5	0.70	NA	-1.87	0.61	-34.21	0.28	-75.29	0.45	-113.33	0.61	0.60	
3	IRMS	NM		2	0.10	NM	-2.44	0.48	-37.58	0.33	-77.51	0.31	-113.07	0.43		
4	LGR	NM	NA	3	NM	NA	-5.88	1.37	-40.48	1.40	-78.10	1.73	-101.90	1.79		
5	IRMS	Y	Pt	1	1000	NM	-2.35	0.86	-38.20	0.68	-78.41	0.86	-114.34	0.85		
6a	IRMS	Y		3	5000	2.00	-3.66	1.50	-38.88	1.10	-79.86	1.60	-115.86	1.90		
6b	PIC	Y	NA		1.75	NA	-3.60	0.6	-38.16	0.20	-77.17	0.30	-101.90	0.50		

Lab ID	Equip.	CA	Method	NS	AW (μL)	ET (hr)	IAEA-OH-13 $\delta^2\text{H}$ 1 σ [‰]		IAEA-OH-14 $\delta^2\text{H}$ 1 σ [‰]		IAEA-OH-15 $\delta^2\text{H}$ 1 σ [‰]		IAEA-OH-16 $\delta^2\text{H}$ 1 σ [‰]		Stated Unc.
Reference values							-2.84	0.60	-38.30	0.36	-78.26	0.38	-114.62	0.43	
7	IRMS	Y		2	3000	1.45	-3.40	2.2	-41.50	2.50	-79.60	2.40	-116.6	1.8	
8	IRMS	N		3	400	1.50	-2.88	0.93	-37.99	-1.04	-78.70	0.86	-115.25	0.96	0.50
9	IRMS	Y			0.70	NM	-2.50	1.08	-37.01	1.11	-75.82	0.95	-113.81	0.82	
11	LGR	NM	NA	6	1000	NA	-2.93	0.35	-38.93	0.48	-78.23	0.35	-114.56	0.15	1.00
13a	IRMS	Y		3	0.50	NM	-1.74	1.16	-36.82	1.14	-77.13	0.96	-115.16	1.13	
15a	IRMS	NM	Cr	3	5000	0.05	-3.63	0.21	-38.53	0.33	-78.17	0.31	-114.10	0.50	
15b	PIC	NM	NA		NM	NA	-3.93	0.243	-38.68	0.37	-78.50	0.36	-114.43	0.37	
19	LGR	N	NA	3	0.97	NA	-1.60	0.34	-37.80	0.46	-78.95	0.60	-116.30	0.53	1.00
22	LGR	N	NA		NM	NA	0.42	1.46	-37.47	2.57	-79.83	1.07	-116.38	0.13	
24	IRMS	Y		3	1.00	NM	4.92	0.34	-39.74	0.22	-79.77	0.33	-114.74	0.39	
25a	PIC	Y	NA	3	2.00	NA	-4.50	0.20	-40.10	0.20	-80.40	0.20	-116.80	0.10	
25b	IRMS	Y		3	NM	18.00	-3.10	0.40	-37.30	0.80	-78.40	0.80	-114.70	1.00	
26	IRMS	NM	Cr	3	2000	0.02	-2.96	0.52	-40.92	0.68	-82.15	0.42	-120.54	0.46	
27a	IRMS	N		3	200	4.00	-3.60	2.10	-40.10	1.60	-79.90	1.90	-114.40	2.50	
27b	IRMS	NM		3	200	4.00	-4.80	3.00	-41.40	1.60	-82.00	2.30	-115.50	2.30	
28	IRMS	Y		2	NM	NM	-2.70	1.20	-39.20	1.20	-79.50	2.00	-116.70	1.50	
29	IRMS	NM			1.00	NM	-2.41	0.74	-38.34	0.81	-78.40	0.82	-115.94	0.72	
30	LGR	NM	NA		NM	NA	-2.85	0.17	-38.44	0.27	-78.60	0.26	-114.98	0.21	
31	PIC	Y	NA		NM	NA	-4.51	1.60	-38.51	1.74	-78.65	2.53	-115.60	1.11	
34	IRMS	Y		3	300	6.00	-2.55	1.67	-38.17	1.75	-78.99	1.75	-115.36	1.99	
35	PIC	NM	NA	2	10.00	NA	-4.40	1.80	-37.30	1.50	-77.10	1.60	-114.10	1.20	1.46
36	PIC	Y	NA		1.00	NA	-4.40	0.67	-38.70	0.40	-78.50	0.25	-114.90	0.27	0.60
37	IRMS	Y	Pyro	2	0.15	NA	-6.60	0.80	-43.90	0.72	-82.70	0.66	-117.80	1.41	2.00
38a	PIC	N	NA	3	1.80	NA	-2.41	0.14	-38.52	0.12	-79.43	0.37	-116.31	0.39	
38b	IRMS	N		3	0.10	NM	-1.01	0.52	-37.52	0.55	-78.45	0.59	-115.81	0.58	
39	IRMS	Y		3	200	34.00	-2.93	0.57	-39.16	0.61	-76.11	2.86	-116.11	0.77	

Lab ID	Equip.	CA	Method	NS	AW (μL)	ET (hr)	IAEA-OH-13 $\delta^2\text{H}$ $1\sigma[\text{‰}]$		IAEA-OH-14 $\delta^2\text{H}$ $1\sigma[\text{‰}]$		IAEA-OH-15 $\delta^2\text{H}$ $1\sigma[\text{‰}]$		IAEA-OH-16 $\delta^2\text{H}$ $1\sigma[\text{‰}]$		Stated Unc.
Reference values							-2.84	0.60	-38.30	0.36	-78.26	0.38	-114.62	0.43	
40	LGR	Y	NA	2	1.00	NA	-3.25	0.75	-38.29	0.68	-78.65	1.82	-115.25	1.37	
41	PIC	Y	NA	2	1.77	NA	-1.70	0.60	-37.20	0.40	-77.90	0.20	-114.80	0.50	0.30
43	PIC	N		2	1.80	NM	-3.58	0.42	-38.27	0.10	-77.78	0.28	-113.96	0.21	0.33
44	IRMS	Y	Cr		0.30	NM	-3.56	0.20	-38.67	0.50	-79.92	0.95	-114.79	0.18	
45	IRMS	N	Cr	7	0.50	NM	-2.77	0.24	-38.37	0.14	-78.07	0.12	-114.82	0.19	
47a	IRMS	NM	Cr	4		NM	-2.04	0.74	-37.49	0.76	-77.60	0.63	-114.02	0.74	0.80
47b	PIC	NM	NA			NA	-2.11	0.30	-36.85	0.40	-77.22	0.39	-114.15	0.43	0.60
50a	IRMS	Y	Pt	3	4000	0.75	-2.45	0.16	-38.77	0.22	-78.47	0.20	-114.82	0.18	0.60
50b	PIC	NM	NA	3	1.70	NA	-1.84	0.27	-37.22	0.25	-77.71	0.20	-113.77	0.12	0.60
51	LGR	N	NA	3	1.00	NA	-2.20	0.40	-39.70	0.50	-78.80	0.20	-115.70	0.30	
52	IRMS	N		3	0.20	48.00	-1.84	0.65	-38.92	0.52	-78.66	0.73	-115.06	0.72	3.00
53a	IRMS	NM	Cr	3	500	NM	-3.20	0.25	-38.60	0.26	-78.50	0.32	-114.50	0.38	0.90
53b	PIC	NM	NA	3	10.00	NA	-2.60	1.49	-37.80	1.07	-77.60	0.84	-113.70	0.52	0.80
56	PIC	NM	NA	3		NA	-2.30	0.30	-37.40	0.30	-77.60	0.50	-114.10	0.50	0.50
57	IRMS	NM	Pyro	2	0.40	20.00	-2.80	0.80	-37.40	0.90	-77.50	1.02	-113.00	0.51	
58	IRMS	Y	Pyro	3		NM	-0.80	0.65	-37.10	1.06	-76.10	0.78	-113.40	0.02	3.00
59	IRMS	Y	Zn	3	8.00	30.00	-4.26	0.57	-40.22	0.50	-80.34	0.48	-116.02	1.00	1.00
60a	IRMS	NM		2	1.00	NM	-2.59	0.28	-37.92	0.66	-78.35	0.45	-114.66	0.40	
60b	PIC	NM	NA	2		NA	-2.82	0.06	-37.87	0.19	-78.18	0.51	-113.63	0.54	
61	IRMS	Y	Cr	2	0.90	NM	-4.30	0.70	-40.20	1.00	-79.50	0.50	-115.80	0.70	
62	IRMS	NM	Pt	1	5000	1.00	-6.08	1.05	-38.25	1.79	-79.72	1.13	-115.23	1.69	
63a	IRMS	N	Cr	4	1.20	NM	-2.13	0.98	-37.74	0.83	-77.91	1.06	-114.16	0.77	
63b	PIC	N	NA	4	1.85	NA	-2.25	1.07	-37.86	0.92	-78.02	0.81	-114.48	0.78	
64a	IRMS	NM	Pyro	2	500	24.00	-5.40	3.90	-41.50	1.20	-78.60	1.80	-115.50	1.40	
64b	IRMS	NM		2	0.70	NM	-4.30	0.90	-38.30	0.90	-78.10	0.40	-115.20	0.80	
66	IRMS	Y	Zn	3	10	NA	-1.30	0.46	-37.76	0.53	-77.23	0.78	-114.05	0.34	

Lab ID	Equip.	CA	Method	NS	AW (μL)	ET (hr)	IAEA-OH-13 $\delta^2\text{H}$ 1 σ [‰]		IAEA-OH-14 $\delta^2\text{H}$ 1 σ [‰]		IAEA-OH-15 $\delta^2\text{H}$ 1 σ [‰]		IAEA-OH-16 $\delta^2\text{H}$ 1 σ [‰]		Stated Unc.
Reference values							-2.84	0.60	-38.30	0.36	-78.26	0.38	-114.62	0.43	
67	LGR	NM	NA	3	1.00	NA	-0.71	0.42	-37.11	0.66	-78.88	0.68	-117.27	0.63	
69	LGR	NM	NA	3	0.75	NA	-4.00	2.00	-40.10	1.80	-78.80	1.50	-114.40	1.80	
70	IRMS	Y	Pyro	3	1.00	NM	-5.80	0.70	-38.90	0.70	-78.70	0.70	-115.30	0.70	0.70
71	LGR	NM	NA			NA	-3.28	0.77	-38.82	0.48	-78.69	0.28	-115.17	0.49	
72	IRMS	Y	Cr	3		NM	-4.70	2.60	-39.80	2.70	-79.80	2.50	-115.80	2.80	
74a	IRMS	NM		3		4.00	-2.10	0.39	-40.10	0.35	-80.80	0.64	-115.00	1.13	1.00
74b	IRMS	NM		3	2.00	NM	-4.00	0.10	-39.20	1.22	-79.90	1.16	-115.70	0.90	2.00
75a	IRMS	N	Cr	3	0.50	NM	-1.30		-37.20	0.30	-77.80	0.00	-115.90	0.40	
75b	LGR	N	NA	3		NA	-1.30	0.10	-37.00	0.10	-78.40	0.00	-115.90	0.50	
78	LGR	NM	NA	3		NA	-3.00	0.90	-38.40	0.70	-79.70	0.60	-116.60	0.60	
79	IRMS	N	EA	2	0.20	NM	-3.70	0.74	-38.60	0.79	-75.40	0.62	-115.00	0.42	
80a	IRMS	Y		2		6.00	-2.86	1.25	-38.44	1.25	-77.86	1.25	-114.34	1.25	
80b	PIC	Y	NA	2	1.80	NA	-2.63	0.78	-38.22	0.78	-78.64	0.78	-115.45	0.78	
81	PIC	Y	NA	2		NA	-1.10	1.10	-38.60	1.20	-77.10	1.00	-115.80	0.50	
82	LGR	N	NA	5	0.75	NA	-3.30	1.00	-38.50	0.90	-78.40	0.40	-114.80	0.60	0.60
84	IRMS	N	Cr	2	1.00	NM	-2.20	0.22	-38.10	0.41	-78.30	0.24	-114.60	0.23	0.30
86a	LGR	Y	NA	2		NA	-3.20	2.00	-37.70	2.00	-77.40	2.00	-113.80	2.00	
87	IRMS	Y	Pt	2		1.00	-1.78	0.79	-38.29	0.69	-77.07	0.92	-114.32	0.91	
88a	IRMS	N	Cr	5		NM	-2.80	0.36	-38.50	0.18	-79.10	0.39	-115.10	0.10	0.80
89	IRMS	N	Pt	3		4.00	-2.80	0.50	-38.80	0.40	-78.30	0.60	-114.60	0.60	1.05
90	IRMS	NM	Cr	2	1.00	NM	-2.20		-38.40		-78.40		-114.20		
91	LGR	N	NA	3		NA	0.19	0.05	-37.92	0.13	-78.09	0.21	-114.82	0.20	
93	IRMS	Y		3	1000	1.00	-2.80	0.60	-38.90	0.20	-78.20	0.30	-114.60	0.50	
95	IRMS	NM	Pyro	1		NM	-3.60	0.77	-37.30	0.60	-75.89	0.87	-111.39	1.03	0.65
96	IRMS	N	Cr	2	0.50	NM	-3.04	0.60	-38.83	0.40	-78.99	0.20	-115.29	0.30	0.70
98	PIC	NM	NA	4	2.00	NA	-2.59	0.11	-38.26	0.13	-78.34	0.11	-114.59	0.11	

Lab ID	Equip.	CA	Method	NS	AW (μL)	ET (hr)	IAEA-OH-13 $\delta^2\text{H}$ 1 σ [‰]		IAEA-OH-14 $\delta^2\text{H}$ 1 σ [‰]		IAEA-OH-15 $\delta^2\text{H}$ 1 σ [‰]		IAEA-OH-16 $\delta^2\text{H}$ 1 σ [‰]		Stated Unc.
Reference values							-2.84	0.60	-38.30	0.36	-78.26	0.38	-114.62	0.43	
99	IRMS	N	Pt	2	2000	1.40	-4.33	0.29	-40.12	0.21	-79.70	0.34	-115.79	0.23	
100	IRMS	N		5	3000	21.00	-3.50	0.35	-38.90	0.17	-78.90	0.45	-115.30	0.22	1.00
101	IRMS	Y		2	200	NM	-2.44	0.88	-38.53	0.58	-78.44	0.60	-114.90	0.80	1.00
102	LGR	N	NA	3		NA	-1.40	0.41	-37.80	0.62	-78.70	0.71	-114.80	0.74	0.92
104	IRMS	Y	Pt	3	4000	2.00	-3.21	0.34	-38.52	0.47	-78.47	0.52	-114.83	0.62	0.50
105	LGR	Y	NA	3		NA	-3.50	0.48	-39.40	0.41	-79.10	0.23	-115.10	0.44	
106	LGR	Y	NA	2		NA	-1.50	0.67	-38.50	0.29	-77.50	0.41	-115.40	0.98	
108	IRMS	NM		3		3.00	-2.91	0.31	-38.66	0.33	-78.94	0.47	-115.16	0.43	
109	LGR	NM	NA	3		NA	-2.21	1.56	-37.88	1.16	-78.29	1.16	-114.53	1.04	
111	IRMS	Y		2	200	24.00	-2.90	2.46	-38.70	4.04	-80.30	3.02	-115.60	2.88	2.00
112a	IRMS	N		3	3000	3.00	-3.59	0.60	-38.85	0.30	-78.36	0.31	-113.86	0.50	
112b	LGR	N	NA	3		NA	-3.12	0.51	-38.75	0.47	-78.38	0.64	-113.87	0.88	
113A	IRMS	Y	Pt	4		1.00	-3.87	0.57	-39.80	0.64	-80.09	0.71	-117.40	0.54	
113b	PIC	Y	NA	4		NA	-2.36	0.42	-38.34	0.37	-79.23	0.39	-116.32	0.35	
114	LGR	Y	NA	3		NA	-3.20	0.70	-39.70	0.70	-80.60	0.70	-117.20	0.80	
116	IRMS	Y		3	4000	1.50	-4.00	1.20	-39.35	1.30	-79.50	1.30	-115.40	1.50	1.00
118a	IRMS	Y	Zn	2	2.00	NM	-0.40	0.73	-34.60	0.87	-76.80	0.70	-113.80	0.46	
118b	PIC	NM	NA	2	NM	NA	-4.10	0.01	-39.40	0.99	-79.30	0.88	-115.60	0.58	
119a	LGR	N	NA	2	NM	NA	-2.40	0.68	-37.40	1.14	-77.84	1.78	-113.40	1.50	
119b	LGR	NM	NA		NM	NA			-37.10	1.20	-77.21	1.04	-114.20	0.50	
119c	LGR	NM	NA		NM	NA	-2.60	0.39	-38.30	0.58	-78.44	0.27	-115.10	0.10	
120	IRMS	NM			NM	NM	-1.92	0.34	-38.08	0.06	-78.58	0.38	-115.69	0.04	
122	LGR	N	NA	3	0.75	NA	-2.10	0.91	-38.74	0.87	-79.51	1.09	-116.16	0.61	
123	LGR	Y	NA	3	1.20	NA	-2.84	0.86	-38.68	0.73	-78.22	0.48	-115.29	0.61	
125	LGR	N	NA	3	NM	NA	-0.29	0.32	-33.69	0.29	-74.82	0.16	-112.97	0.48	1.00
126	IRMS	N		3	200	1.50	-0.90	2.00	-36.90	1.00	-76.50	2.10	-113.20	1.30	

Lab ID	Equip.	CA	Method	NS	AW (μL)	ET (hr)	IAEA-OH-13 $\delta^2\text{H}$ 1 σ [‰]		IAEA-OH-14 $\delta^2\text{H}$ 1 σ [‰]		IAEA-OH-15 $\delta^2\text{H}$ 1 σ [‰]		IAEA-OH-16 $\delta^2\text{H}$ 1 σ [‰]		Stated Unc.
Reference values							-2.84	0.60	-38.30	0.36	-78.26	0.38	-114.62	0.43	
127	IRMS	Y		2	3000	3.00	-2.01	0.58	-38.64	0.77	-78.41	0.65	-114.66	0.33	
129	LGR	NM	NA	3	NM	NA	-3.41	0.31	-38.42	0.32	-78.29	0.40	-114.93	0.64	1.00
130	LGR	NM	NA	3	NM	NA	-2.59	0.39	-39.35	0.74	-79.22	1.08	-117.16	0.49	1.00
132	LGR	N	NA	3	1.20	NA	-2.84	1.22	-37.97	0.97	-78.07	0.79	-114.73	0.83	
133a	IRMS	N	Pyro	3	0.10	NM	-1.20	0.73	-39.20	0.45	-80.80	0.87	-119.50	0.44	
133b	IRMS	NM			NM	NM	-1.20	0.38	-39.70	0.43	-79.60	0.33	-116.40	0.36	
135a	PIC	Y	NA	3	2.00	NA	-2.7	0.15	-38.3	0.02	-78.4	0.15	-114.7	0.27	
135b	PIC	Y	NA	3	2.00	NA	-2.8	0.09	-38.4	0.02	-78.4	0.15	-114.7	0.53	
136	IRMS	N		2	2000	3.00	-2.50	0.44	-38.30	0.43	-78.20	0.30	-114.30	0.22	
137	LGR	N	NA	2	1.20	NA	-2.41	0.96	-37.13	0.96	-76.38	0.79	-113.14	0.47	
138	PIC	NM	NA		NM	NA	-1.90	0.20	-37.30	0.20	-77.40	0.30	-113.90	0.50	
139	IRMS	NM		2	NM	1.00	-1.70	0.06	37.60	0.15	-78.30	0.26	-115.50	0.44	
140b	PIC	NM	NA	2	1.00	NA	-2.70	0.08	-39.10	0.43	-78.60	0.22	-114.50	0.39	2.00
141	IRMS	N	Pt	2	200	12.00	-3.60	1.43	-39.20	1.22	-78.50	0.73	-115.30	0.98	1.50
142	LGR	N	NA	2	NM	NA	-3.50	1.10	-38.41	0.67	-78.12	0.61	-115.00	1.39	
143	LGR	Y	NA	3	NM	NA	-2.40	0.17	-38.63	0.24	-80.32	0.21	-118.40	0.29	
144	IRMS	N		3	1000	9.00	-2.04	0.32	-37.83	0.30	-77.64	0.49	-113.89	0.27	
145	LGR	NM	NA	7	1.00	NA	-2.90	0.73	-39.80	0.70	-80.30	0.40	-117.20	0.20	
146a	IRMS	N		1	3000	1.00	-1.92	0.61	-37.79	0.48	-78.34	0.45	-114.83	0.66	1.00
146b	PIC	N	NA	3	1.00	NA	-2.27	0.11	-38.10	0.17	-78.27	0.16	-115.09	0.20	1.00
146c	PIC	N	NA	3	1.00	NA	-2.49	0.24	-37.83	0.38	-78.44	0.51	-114.71	0.51	1.00
147a	IRMS	N	Pt	3	5000	2.00	-2.43	0.52	-38.20	0.43	-78.15	0.52	-114.75	0.43	
147b	PIC	N	NA	3	2000	NA	-2.65	0.82	-37.88	0.52	-78.45	0.72	-114.83	0.72	
148	IRMS	N	Cr	2	1.50	NM	-9.40	0.85	-41.25	0.49	-78.90	0.14	-113.35	0.78	1.00
149a	PIC	Y	NA	3	1.80	NA	-2.30	0.20	-37.80	0.20	-77.90	0.30	-114.30	0.40	
152	IRMS	NM	Pt	3	1000	1.00	-2.10	0.40	-38.20	1.00	-78.90	1.10	-114.80	0.60	

Lab ID	Equip.	CA	Method	NS	AW (μ L)	ET (hr)	IAEA-OH-13 $\delta^2\text{H}$ 1 σ [‰]		IAEA-OH-14 $\delta^2\text{H}$ 1 σ [‰]		IAEA-OH-15 $\delta^2\text{H}$ 1 σ [‰]		IAEA-OH-16 $\delta^2\text{H}$ 1 σ [‰]		Stated Unc.	
Reference values							-2.84	0.60	-38.30	0.36	-78.26	0.38	-114.62	0.43		
153	IRMS	Y		3	400	2.50	-3.04	0.23	-38.40	0.19	-78.30	0.18	-114.50	0.34		
154	IRMS	NM	Pt	3	200	1.00	-3.33	0.60	-39.46	1.00	-79.70	1.00	-115.95	1.40		
155a	IRMS	Y	Pt	3	4000	NM	-2.3	0.9	-37.7	0.8	-78.0	0.8	-113.8	0.9	0.21	
155b	PIC	Y	NA	3	1.00	NA	-4.17	1.16	-37.69	0.55	-77.68	0.42	-114.39	0.32		
155c	LGR	Y	NA	3	0.75	NA	-3.99	0.64	-38.68	0.55	-78.78	0.34	-115.34	0.55		
155d	LGR	Y	NA	3	0.75	NA	-3.96	0.88	-38.79	0.90	-79.11	1.29	-115.50	0.65		
155e	LGR	Y	NA	3	0.75	NA	-3.64	1.87	-39.00	2.39	-77.82	1.62	-114.22	1.42		
157	IRMS	Y	Zn	1	NM	1.00	-0.10	1.02	-37.30	0.45	-79.70	0.38	-116.00	0.25	1.02	
159	IRMS	Y	Cr	2	NM	0.10	-0.58	1.02	-35.86	0.17	-77.28	0.69	-113.97	0.42		
160	PIC	Y	NA	10	1.00	NA	-2.40	0.20	-38.50	0.50	-77.90	0.10	-114.70	0.20		
161	LGR	Y	NA	3	NM	NA	-0.07	0.02	-35.26	0.10	-75.64	0.12	-112.35	0.09	0.60	
162	IRMS	N	Cr	3	0.20	NM	-2.78	0.44	-38.61	0.66	-78.64	0.31	-115.56	0.44	0.80	
165	PIC	NM	NA	3	NM	NA	-2.61	1.79	-37.01	0.53	-77.82	0.83	-116.17	0.89		
168	IRMS	Y	Pyro	4	NM	NM	-1.40	0.50	-37.30	0.60	-77.30	0.60	-113.40	0.60		
170	PIC	N	NA	3	NM	NA	-3.28	0.65	-38.66	0.54	-78.31	0.15	-114.98	0.50		
172a	PIC	Y	NA	3	1.70	NA	-2.61	0.31	-44.33	1.72	-78.05	0.83	-116.38	0.28		
175	LGR	NM	NA	3	1.20	NA	-12.90	1.60	-46.30	0.90	-85.90	0.70	-122.70	0.70		
177	LGR	NM	NA	3	NM	NA	-3.82	0.75	-38.76	0.55	-79.17	0.40	-116.30	0.86		
178	PIC	NM	NA	3	NM	NA	-2.07	0.33	-37.29	0.32	-77.54	0.31	-114.21	0.42		
180	PIC	NM	NA	no!	1.10	NA	-7.13	0.67	-42.18	0.41	-82.77	0.26	-119.38	0.30		Manual injection
181	LGR	NM	NA	3	NM	NA	1.08	0.42	-35.50	1.08	-77.05	0.19	-114.30	0.56	1.00	
182a	IRMS	Y		2	5.00	35.00	-1.84	0.50	-37.79	0.60	-76.62	0.50	-114.71	0.40	1.00	
182b	LGR	Y	NA	2	0.75	NA	-1.54	0.40	-36.56	0.30	-76.59	0.30	-113.58	0.60	1.00	
183	PIC	Y	NA		1.80	NA	-1.80	0.08	-37.30	0.30	-78.00	0.29	-115.03	0.33		

APPENDIX II. LIST OF PARTICIPANTS

Note: This list is sorted alphabetically by country and is not related to the assigned Laboratory ID number used throughout the report.

Iris Bakiri
Center of the Applied Nuclear Physics
str. Thoma Filipeu
P.O.Box 85
Qesarake,Tirana
Albania
++355 2451371
bakiriiris@yahoo.com

Water Research Laboratory
School of Civil and Environmental Engineering
University of New South Wales
110 King St
2093 Manly Vale
Australia
++61 2 8071 9800
m.andersen@wrl.unsw.edu.au

Hector Panarello
Instituto de Geocronologia y Geologia
Isotopica (INGEIS)
Pabellon INGEIS, Ciudad Universitaria
1428 Buenos Aires
Argentina
++ 54 (011) 4783-3021/23
panarello@ingeis.uba.ar

Grzegorz Skrzypek
West Australian Biogeochemistry Centre
School of Plant Biology
Reception Desk at Botany Building
UWA Campus, Hackett Drive Entrance 3
6009 Crawley
Australia
++61 8 6488 4584
grzegorz.skrzypek@uwa.edu.au

Daniel Martinez
Centro de Geología de Costas y del
Cuaternario
Universidad Nacional de Mar del Plata
Calle Funes 3350
7600 Mar del Plata
Argentina
++54 23 4754060
demarti@mdp.edu.ar

Manzoor Choudhry
IAEA
P.O. Box 200
A-1400 Vienna
Austria
+43-1-2600-21758
M.CHOUDHRY@iaea.org

Fred Leaney
CSIRO Land and Water
Waite Rd. Gate 5.
5064 Urrbrae
Australia
++61(-8)-8303 8719
fred.leaney@csiro.au

Albrecht Leis
Joanneum Research Forschungsgesellschaft
GmbH
Inst. of Hydrogeology and Geothermics
Elisabethstrasse 16/II
8010 Graz
Austria
++43 (316) 876-1485
albrecht.leis@joanneum.at

Barbara Neklapilova
Australian Nuclear Science and Technology
Organisation
Locked Bag 2001
New Illawarra Rd
2234 Lucas Heights
Australia
++61 2 9717 3987
barbara.neklapilova@ansto.gov.au

Stefan Wyhlidal
Austrian Institute of Technology
Konrad Lorenz Strasse 24
3430 Tulln
Austria
++43 (0) 50550-3615 |
stefan.wyhlidal@ait.ac.at

Martin Andersen

Christoph Spötl
Institut für Geologie und Paläontologie

Universität Innsbruck
Innrain 52
6020 Innsbruck
Austria
+43 (512) 507 5593
Christoph.Spoetl@uibk.ac.at

Marcelo Moreira
CENA / Univ S. Paulo
Av. CENTENARIO, 303
13416-903 Piracicaba
Brazil
+55 (19) 3429 4075
mmoreira@cena.usp.br

José Godoy
Instituto de Radioproteção e Dosimetria
Avenida Salvador Allende s/n
Barra da Tijuca
CEP 22180 Rio de Janeiro
Brazil
55-21-2442-1927
jmgodoy@ird.gov.br

Paul Eby
Alberta Innovates Technology Futures
#3-4476 Markham St.
V8Z 7X8 Victoria
Canada
++250-483-3290
peby@uvic.ca

Jean-Francois Helie
Universite du Quebec a Montreal
201 Avenue Président-Kennedy
7th floor, room PK-7150
H2X 3Y7 Montréal
Canada
++1 514-987-3000 #2413
helie.jean-francois@uqam.ca

Alvin Kwan
Biogeochemical Analytical Laboratory
2-255, CCIS Building,
University of Alberta,
T6G 2E9 Edmonton
Canada
++780 492-9034
alvin.kwan@ualberta.ca
Stephen Taylor
University of Calgary, Dept of Physics

2500 University Dr. N.W.
T2N 1N4 Calgary
Canada
++1 403 220 8268
steve.taylor@ucalgary.ca

Robert Drimmie
Isotope Tracer Technologies
85 Bathurst Drive, Unit D
N2V 1N2 Waterloo
Canada
++(519) 888-4567 Ext. 32580
bob@it2isotopes.com

Richard Heemskerk
UwEILAB
C/O Chemistry Stores
University of Waterloo
200 University Ave. W.
N2L 3G1 Waterloo
Canada
++519-888-456735838
rkhmskrk@uwaterloo.ca

Geoff Koehler
NHRC Stable Isotope Hydrology and Ecology
Laboratory
Environment Canada
11 Innovation Blvd.
S7N 3H5 Saskatoon, SK
Canada
++1 306 975 5778
Geoff.Koehler@EC.GC.CA

Evelyn Aguirre
Laboratorio de Isotopos Ambientales
Comision Chilena de Energía Nuclear
Cen La Reina
Nueva Bilbao 12501, LC, Santiago
Chile
++56 (2) 3646113
eaguirre@cchen.cl

Felipe Rodríguez
Servicio Nacional de Geología y Minería
TilTil 1993, Ñuñoa.
Santiago
Chile
++56 (2) 2385292
fllona@sernageomin.cl

Zhonghe Pang
Institute of Geology and Geophysics
Chinese Academy of Sciences,
P.O.Box 9825
19 Beitucheng Western Road,
Chaoyang District
100029 Beijing
China
++86-10-82998613
pangtravel@hotmail.com

Lin Zhang
Laboratory of Groundwater Science
Institute of Hydrogeology and Environmental
Geology ,
Chinese Academy of Geosciences
92,Zhongshan East Road
050803 Zhengding
China
++311-88012330
zh15369@163.com

Tulio Ruiz-Alvarez
Laboratorio de Calidad Ambiental
"Morrosquillo"
Subdirección de Gestión Ambiental
CARSUCRECr. 25 No. 25-101 Av. Ocala
Sincelejo-Sucre
Colombia
++57 5 2749994 Ext.115
tuliorrui@yaho.com

Wilson Beita-Sandi
Servicio Nacional de Aguas Subterráneas
Riego y Avenamiento
Ciudad Universitaria Rodrigo Facio,
San Pedro de Montes de Oca
Barrio Cuba
506 San José
Costa Rica
++00506 2511 3323
wilson.beita@ucr.ac.cr

Zvezdana Roller-Lutz
Medical Faculty, Rijeka University
Stable Isotope Laboratory
Brace Branchetta 20
51 000 Rijeka
Croatia
++385 51 651 210
roller@medri.hr

Martin Sanda
Czech Technical University in Prague
Faculty of Civil Engineering
K143 Thakurova 7
166 29 Praha 6
Czech Republic
++420 224353739
martin.sanda@fsv.cvut.cz

Roger Bautista-Murcia
ANDA, Laboratorio de Control de Calidad
Boulevard del Hipódromo #609
Colonia San Benito,
El Salvador
++5032247 2533
rbautista@anda.gob.sv

Tonu Martma
Institute of Geology at Tallinn University of
Technology
Ehitajate tee 5
19086 Tallinn
Estonia
++372 620 30 10
martma@gi.ee

Tenalem Ayenew
Addis Ababa University
Department of Geology and Geophysics
P.O. Box 2380
Addis Ababa
Ethiopia
++251 1 236294
Tenalema@geol.aau.edi.et

Nina Kortelainen
Laboratory for Isotope Geology
Geological Survey of Finland GTK
PO Box 96
02151 Espoo
Finland
++358 20 550 2305
nina.kortelainen@gtk.fi

Alexandre Zahariev
CNRS - IPHC - DEPE
23, rue Becquerel
67087 Strasbourg cedex 2
France
++03 88 10 69 42

alexandre.zahariev@iphc.cnrs.fr

Christine Flehoc
BRGM - ANA/ISO
3 avenue C. Guillemin
BP 6009
45060 Orléans
France
++33 2 38 64 34 13
c.flehoc@brgm.fr

Gael Monvoisin
Laboratoire Interactions et Dynamiques des
Environnements de Surface - UMR
8148 Bâtiment 504 - 3^e étage - Porte 326
Rue du belvédère
91405 Orsay
France
++331 69 15 71 74
gael.monvoisin@u-psud.fr

Christine Vallet-Coulomb
UPCAM - CEREGE
BP 80 - Europole Mediterranéen de l'Arbois
13545 Aix-en-Provence cedex 4
France
++33 4 42971593
vallet@cerege.fr

Sonia Falourd
LSCE
CEA SACLAY
Bat 703 91191
91191 Gif sur Yvette
France
++33169083877
falourd@cea.fr

Nicolas Patris
LAMA - HSM UMR5569 - IRD
Laboratoire d'Analyse des isotopes de l'eau
Maison des Sciences de l'Eau
Université Montpellier 2, CC 57
34095 Montpellier Cedex 5
France
++33 4 67 14 90 31
npatris@msem.univ-montp2.fr

Marina Gillon
Université d'Avignon et des pays de Vaucluse
Laboratoire d'hydrogéologie

33 rue Louis Pasteur
84000 Avignon
France
++33 4 90 14 44 67
marina.gillon@univ-avignon.fr

Freddy Thomas
Eurofins
Site de la Géraudière
Rue Pierre Adolphe Bobierre
B.P. 42301
44323 Nantes
France
++33 2-51-83-21-00

Kay Knoeller
Helmholtz Centre for Environmental Research
Department Catchment Hydrology
Theodor-Lieser-Str. 4
06120 Halle
Germany
++49-345-558-5433, -5240
kay.knoeller@ufz.de

Matthias Gehre
Laboratory for stable Isotopes (LSI)
Department for Isotope Biogeochemistry
Helmholtz-Centre for Environmental Research
UFZ Permoserstrasse 15
04318 Leipzig
Germany
++49 (0341) 235 2252
matthias.gehre@ufz.de

Paul Koeniger
Leibniz Institute for Applied Geosciences
Geochronology and Isotope Hydrology (S3)
Stilleweg 2
30655 Hannover
Germany
++49 (511) 6432527
Paul.Koeniger@liag-hannover.de

Willy Brandt
Stable Isotope Laboratory
Max-Planck-Institute for Biogeochemistry
Winzlaer Str. 10, P.O.Box 100164
07701 Jena
Germany
++49 (-3641) -643718
wbrandt@bgc-jena.mpg.de

Hilmar Förstel
TÜV Rheinland Agroisolab GmbH
Prof.-Rehm-Strasse 6
52428 Jülich
Germany
++49 (0) 2461 93134-0
h.foerstel@agroisolab.de

Hanno Meyer
Potsdam AWI
Stable Isotope Laboratory
Alfred Wegener Institute for Polar and Marine
Research Research Unit Potsdam
Telegrafenberg A 43
14473 Potsdam
Germany
++49 (-331) -288-2115
Hanno.Meyer@awi.de

Barbara Herbstritt
Institut für Hydrologie
Universität Freiburg
Fahnenbergplatz
79098 Freiburg
Germany
++49- (0) 761-203-3539
barbara.herbstritt@hydrology.uni-freiburg.de

Andrey Voropaev
Hydroisotop
Woelkestr. 9
85301 Schweitenkirchen
Germany
++49 8444 92890
av@hydroisotop.de

Edward Komla Pappah Bam
Isotope Hydrology Laboratory, GAEC
C/O Resident Representative
UN Development Programme in Ghana
P.O. Box 1423
Ring Road Dual Carriage (near Police
Headquarters)
Legon Accra
Ghana
++233 244 753 061
evangelbam@yahoo.com

István Fórizs
Institute for Geochemical Research

Budaörsi út 45.
1112 Budapest
Hungary
++36 1 319 3137
forizs@geochem.hu

Lazslo Palcsu
Institute of Nuclear Research
Hungarian Academy of Sciences
Bem Square 18/c
4026 Debrecen
Hungary
++36 52 509 200
palcsu@atomki.hu

Supriyo Chakraborty
Indian Institute of Tropical Meteorology
Center for Climate Change Research
Dr. Homi Bhabha Road
411008 Pashan-Pune
India
++91-20-2590 4200
supriyo@tropmet.res.in

Bhishm Kumar
Hydrological Investigations Division,
National Institute of Hydrology
Uttarakhand
247667 Roorkee
India
++91 1332 249202
bk@nih.ernet.in

Rashmikant Jani
Geosciences Division
Physical Research Laboratory
(Dept. of Space, Govt. of India)
380 009 Ahmedabad, Gujarat
India
++91-79-26314269
rajani@prl.res.in

Djiono Djiono
Isotope Hydrology and Geothermal
Laboratory,
PAIR-BATAN
Centre for Application of Isotope and
Radiation TechnologyHydrology lab
Jl. Lebak Bulus Raya
12070 Jakarta
Indonesia

++62 21-7659376
djionobatan@yahoo.com

Vittorio Barella
Iso4
c/o Dipartimento di Scienze della Terra
Via Valperga Caluso 35
10125 Torino
Italy
++39011655027
iso4.dst@unito.it

Paola Iacumin
Parma University
Department of Earth Sciences
Laboratory of Isotope Geochemistry
Parco Area delle Scienze 157/A
43100 Parma
Italy
++39 (521) 905321
paola.iacumin@unipr.it

Barbara Stenni
Dipartimento di Scienze Geologiche,
Ambientali e Marine
Laboratorio di Geochimica Isotopica
Via E. Weiss 2
34127 Trieste
Italy
++39-040-5582153/4
stenni@univ.trieste.it

Carmine Lubritto
Centre for Isotopic Research
Seconda Università degli Studi di Napoli
via Vivaldi, 43
81100 Caserta
Italy
++39 0823 274605
carmine.lubritto@unina2.it

Daniele Penna
Dipartimento Territorio e Sistemi Agroforestali
(TESAF)
Università di Padova
laboratorio IDEA, 2° piano
viale dell'Università 16
35020 Legnaro
Italy
++39.049.8272700
daniele.penna@unipd.it

Fausto Grassa
INGV Istituto Nazionale di Geofisica e
Vulcanologia Sezione di Palermo
Via Ugo La Malfa, 153
90146 Palermo
Italy
++39 0916809406
grassa@pa.ingv.it

Kumiko Goto-Azuma
National Institute of Polar Research
Tokyo 190-8518
Tachikawa
Japan
kumiko@nipr.ac.jp

Takayuki Kuramoto
National Institute of Polar Research
10-3 Midori-cho, Tachikawa
190-8518 Tokyo
Japan
++81 42-512-0758
kuramoto.takayuki@nipr.ac.jp

Kei Yoshimura
IIS Tokyo University
4-6-1 Komaba,
153-8505 Meguro-ku, Tokyo
Japan
++813 -5452-6381
keiyoshi08@gmail.com

Urumu Tsunogai
Earth and Planetary System Science,
Faculty of Science, Hokkaido University
060-0810 Sapporo
Japan
++81-11-706-3586
urumu@mail.sci.hokudai.ac.jp

Atsuko Sugimoto
Faculty of Environmental Earth Science
Hokkaido University
Nishi 5, Kita 10
060-0810 Kita-ku
Japan
++81-11-706-2233
atsukos@ees.hokudai.ac.jp

Hidehiko Kawanishi

SI science co ltd
Kitakatsushikagun
Saitama
345-0023
Japan
++81-480-37-1555
nkawa@shoko.co.jp

Abe Osamu
Hydrospheric Atmospheric Research Center
Nagoya University
4648601 Nagoya University
Japan
++81-52-789-3470
osamu.abe@nagoya-u.jp

Ryu Uemura
Department of Chemistry, Biology and Marine
Science
University of the Ryukyus
903-0213 Senbaru, Nishihara
Japan
++81-98-895-8526
ruemura@sci.u-ryukyu.ac.jp

Ismail Abdeldin
Water Authority of Jordan
Ministry of Water and Irrigation
Khansa'a Street, Biader Wadi Esseer (next to
civil defense department)
11183 Amman
Jordan
++962 (6) 58643761 / 2
ismailabdeldin@yahoo.com

Harish Bhandary
Kuwait Institute for Scientific Research (KISR)
P.O. Box 24885
13109 Kuwait
Kuwait
++965 97155058
hkaup@kISR.edu.kw

Francois Barnich
Centre de Recherche Public - Gabriel
Lippmann
Environment and Biotechnologies Research
Unit
41 rue du Brill
4422 Belvaux
Luxembourg

++352 470 261 806
barnich@lippmann.lu

Roslanzairi Mostapa
Malaysia Nuclear Agency
43000 Bangi, Kajang
Malaysia
++60 3 89250510
roslanzairi@nuclearmalaysia.gov.my

Fareed Vacil
National Environmental Laboratory
National Laboratories Complex
REDUIT
Mauritius
++230 4661080
maurinel@orange.mu

Pedro Morales-Puentes
Laboratorio de Espectrometria de Masas de
Isotopos Estables
Departamento de Geoquimica
Universidad Nacional Autonoma de Mexico
Apartado Postal 70-296,
Circuito de la Investigación s/n Ciudad
Universitaria
04510 Mexico
Mexico
++52 52 55 56 22 43 25
mopuente@servidor.unam.mx

Hamid Marah
Centre national de l' energie, des sciences
Unite Eau et Climat,
BP 1382
10001 Rabat, Agdal
Morocco
++212 37 81 97 57
h_marah@yahoo.com

Anita Aerts-Bijma
Centre for Isotope Research
University of Groningen
Nijenborgh 4
9749 AG Groningen
Netherlands
++31-503634760
A.T.Aerts-Bijma@rug.nl

Karyne Rogers
Institute of Geological & Nuclear Sciences

Stable Isotope Laboratory
PO Box 31-312
30 Gracefield Road
Lower Hut
New Zealand
++64 4 5704637
k.rogers@gns.cri.nz

Robert Van Hale
Department of Chemistry
University of Otago
Union Place West
P.O. Box 56
9054 Dunedin
New Zealand
++64 3 4797915
robertv@chemistry.otago.ac.nz

Ingar Johansen
Institute for Energy Technology
PO Box 40
2027 Kjeller
Norway
++47 481 56 929
ingar@ife.no

Zahid Latif
Isotope Application Division
Pakistan Institute of Nuclear Science &
Technology (PINSTECH),
P.O. Nilore,
45650 Islamabad
Pakistan
++92 51 2208038
zlatifm@yahoo.com

Jorge Condori Ccari
Instituto Peruano de Energía Nuclear (IPEN)
Avenida Canadá 1470
Apartado 1687
Lima 1687
Peru
++51 1 4885050
jcondori@ipen.gob.pe

Guima Urbino
PNO
Chemistry and Isotope Laboratory Section
Energy Development Corporation
Energy Complex, Merrit Road, Fort Bonifacio
1201 Taguig City

Philippines
++63 2 479-9400
urbino.ga@energy.com.ph

Zbigniew Gorczyca
Environmental Physics Group
Faculty of Physics and Applied Computer
Science
ul. Reymonta 19,(room 13)
30-059 Krakow
Poland
++48 12 6173010
gorczyca@novell.ftj.agh.edu.pl

Stanislaw Halas
Mass Spectrometry Laboratory
Institute of Physics
Maria Curie-Skłodowska University
20-031 Lublin
Poland
++48 81 537 6275
halas@tytan.umcs.lublin.pl

Adam Porowski
Institute of Geological Sciences
Stable Isotope Laboratory
Polish Academy of Science
Twarda 51/55,
Poland
++48 (22) 697-87-56
adamp@twarda.pan.pl

Paula Carreira
ITN - Instituto Tecnológico e Nuclear
Departamento de Química
Estrada Nacional nº 10
2686- 953 Sacavém
Portugal
++351 219946179
carreira@itn.pt

Diana Costinel
Physical Chemical, Environmental and Food
Quality Analysis Laboratory
National R&D Institute of Cryogenics and
Isotopic Technologies – ICSI Rm. Valcea,
research & Development Department
Str. Uzinei, nr. 4
240050 Valcea
Romania
++40.250.733890

diana@icsi.ro

Alexey Ekaykin
Climate and Environmental Research
Laboratory
Arctic and Antarctic Research Institute,
38 Beringa st.
199397 St. Petersburg
Russia
++7 812 337 31 62
ekaykin@aari.nw.ru

Elena Dubinina
Institute of ore deposits, petrography,
mineralogy and geochemistry of Russian
Academy of Sciences (IGEM RAS),
Staromonetnyi pereulok, 35
119017 Moscow
Russia
++7 (495) 951-15-87
elenadelta@gmail.com

Nada Miljevic
Jaroslav Cerni Institute for Development of
Water Resources
Jaroslava Cernog 80
11226 Belgrade
Serbia
++381 11 390 83 88
emiljevi@vinca.rs

Sonja Lojen
Department of Environmental Sciences
Jožef Stefan Institute
Jamova 39
1000 Ljubljana
Slovenia
++386 (1) 5885 393
sonja.lojen@ijs.si

Stephan Woodborne
CSIR NRE Environmental Analytical
Laboratories
Room 067, Building 4E
Meiring Naude Road
0087 Brummeria
South Africa
++27 2712 841-3402
swoodbor@csir.co.za

Chris Harris

Dept of Geological Sciences
University of Cape Town
13, University Avenue
7701 Rondebosch
South Africa
++27 21 6504886
Chris.Harris@uct.ac.za

Luz Robredo
Unidad de Barreras de Ingeniería y Geológica
Departamento de Medio Ambiente
Av. Complutense 22
Edificio 19
28040 Madrid
Spain
++34 913460834
l.robredo@ciemat.es

Ramón Redondo-Ortega
Laboratorio de Isótopos Estables. SIdI
Facultad de Ciencias, C-IX
Universidad Autónoma de Madrid
28049 Madrid
Spain
++34 914974898
ramon.redondo@uam.es

Fe Díaz Teijeiro
Centro de Estudios y Experimentación de
Obras Públicas (CEDEX)
Ministerio de Fomento
C/ Alfonso XII, 3 y 5
28014 Madrid
Spain
++34 (91) 335 72 64
marife.diaz@cedex.es

María Grille
University Coruña SAI
Scientific Research Support Services
Edificio Servicios Centrais de Investigación
Campus de Elviña s/n
15071-A Coruña
Spain
++981 167 000 ext. 2074
saiutia@udc.es

Pilar Teixidor
University of Barcelona
Serveis Científicotecnics
C/Lluis Sole sabaris, 1-3

08028 Barcelona
Spain
++34 (93) 4021694
teixidor@giga.sct.ub.es

Viraj Edirisinghe
Isotope Hydrology Laboratory
Atomic Energy Authority
Industrial Applications Division
60/460, Baseline Road Orugodawatta
Wellampitiya
Sri Lanka
++94 11 2533427
viraj@aea.ac.lk

Heike Siegmund
Department of Geological Sciences
Stable Isotope Laboratory (SIL)
Stockholm University
Svante Arrhenius Väg 8
106 91 Stockholm
Sweden
++46-8-6747726
heike.siegmund@geo.su.se

Mathias Saurer
Paul Scherrer Institute
5232 Villigen
Switzerland
++41 (56) 310 27 49
matthias.saurer@psi.ch

Benjamin Fischer
Hydrology and Climate
Department of Geography
University of Zurich-Irchel
Room Y25K78
Winterthurerstr. 190
CH-8057 Zürich
Switzerland
++1- (0) 4463 55227
benjamin.fischer@geo.uzh.ch

Michael Berg
Institute of Aquatic Science and Technology
Eawag, Contaminant Hydrology
Swiss Federal Institute of Aquatic Science and
Technology
8600 Duebendorf
Switzerland
++41-44-823 5078 / or 5477

michael.berg@eawag.ch

Markus Leuenberger
Climate and Environmental Physics Lab.
Physics Institute, Universität Bern
Sidlerstrasse 5
3012 Bern
Switzerland
++41 631 44 70
leuenberger@climate.unibe.ch

Chen-Feng You
Isotope Geochemistry Laboratory
Earth Dynamic System Research Center
National Cheng Kung University
No.1, University Road,
701 Tainan City
Taiwan
++886 6-275-7575 ext 65432
cfy20@mail.ncku.edu.tw

Chung-Ho Wang
Stable Isotope Laboratory
Institute of Earth Sciences
Academia Sinica
No.128, Section II, Yen-Chiu-Yuan Road,
Nankang, Taipei
11529 Taipeh
Taiwan
++886 2 2783-9910 ext.513
chwang@earth.sinica.edu.tw

Wutichai Chantarachot
Thailand Institute of Nuclear Technology
(TINT)
9/9 Moo 7, Sai Mun, Ongkharak
26120 Nakorn Nayok
Thailand
++6637-392901 (to 6)#1052
yhangwan@hotmail.com

Rim TRABELSI
Sfax University
Director of Laboratoire Radio-analyses et
Environnement
Université de Sfax
École nationale d'ingénieurs de Sfax
Département de génie géologique
B.P. W
3038 Sfax
Tunisia

++216 74 274 088
rimaenis@yahoo.fr

Alime Temel Dilaver
Ankara Hydrolic Works DSI
Isotope Laboratory
DSİ TAKK Dairesi Başkanlığı
Inönü Bulvarı
06100 Yücepete, Ankara
Turkey
++312 399 27 92
izotop@dsi.gov.tr

Mehmet Ekmekekci
Hacettepe University Karst Center
Stable Isotopes Laboratory
Hacettepe University (Hu-Ukam)
06800 Ankara
Turkey
++990 312 2976725
ekmekci@hacettepe.edu.tr

Barry Thornton
The James Hutton Institute, Stable Isotopes
Section
AB15 8QH Craigiebuckler, Aberdeen
United Kingdom
++44 1224 395 123
b.thornton@macaulay.ac.uk

Carol Arrowsmith
NERC Isotope Geosciences Laboratory
British Geological Survey
Keyworth
NG12 5GG Nottingham
United Kingdom
+44 (0) 115 936 3514
caar@nigl.nerc.ac.uk

George Darling
British Geological Survey
Maclean Building
OX10 8BB Wallingford
United Kingdom
01491 838800
wgd@bgs.ac.uk

Stuart Black
Department of Archaeology
School of Human and Environmental Sciences,
University of Reading,

Whiteknights,
PO Box 227,
RG6 6AB Reading
United Kingdom
(0) 118 3788946
s.black@reading.ac.uk

Jason Newton
NERC Life Sciences
Scottish Universities Environmental Research
Centre
G75 0QF Rankine Avenue
United Kingdom
+44 (0) 1355 270138
Jason.Newton@glasgow.ac.uk

Ian Boomer
School of Geography, Earth & Environmental
Sciences
Stable Isotope Laboratory
Edgbaston
B15 2TT Birmingham
United Kingdom
0121 41 45536
i.boomer@bham.ac.uk

Wolfram Meier-Augenstein
Scottish Crop Research Institute
Stable Isotope Laboratory
The James Hutton Institute
Invergowrie
DD2 5DA Dundee
United Kingdom
+44 (0) 844 928 5428
wolfram.meier-augenstein@scri.ac.uk

Steven Brookes
Iso-Analytical Ltd
The Quantum, Phase 3
Marshfield Bank Business Park
CW2 8UY Crewe
United Kingdom
+44 (0) 1270 509533
steve@iso-analytical.com

Jonathan Wynn
USF Geology Stable Isotope Lab
Department of Geology
University of South Florida
4202 E. Fowler Avenue, SCA 528
33620-8100 Tampa

United States of America
+1 (813) 974-2654 or 9369
jwynn@cas.usf.edu

H. Feng
Kerr Lab (USEPA)
Stable Isotope Lab
919 Research Drive
74820 Ada
United States of America
(580) 436-8676
lu.feng@epa.gov

Steven Pelphey
Isotech Laboratories, Inc.
1308 Parkland Court
61821 Champaign
USA
+1 217-398-3490
steve@isotechlabs.com

Daniel Snow
University of Nebraska
202 Water Sciences Laboratory
68583-0844 Lincoln
USA
(402) 472-9599
dsnow1@unl.edu

Nabil Saad
Picarro, Inc.
3105 Patrick Henry Drive
95054 Santa Clara
USA
1-408-962-3931
nsaad@picarro.com

Doug Baer
Los Gatos Research
67 East Evelyn Avenue, Suite 3
94041-1529 Mountain View
USA
650-965-7772
d.baer@LGRinc.com

Roland Thurston
U.S. Geological Survey
973 Denver Federal Center
80225 Denver
USA
(303) 236-1920

rthurston@usgs.gov

Randy Culp
The University of Georgia
Center for Applied Isotope Studies
120 Riverbend Rd.
30602 Athens
USA
706-542-6122
Rculp@uga.edu

Brian Popp
Hawaii University
Department of Geology and Geophysics
1680 East-West Road
University of Hawaii
96822 Honolulu
USA
(808) 956-6206
popp@hawaii.edu

Dan Reuss
Colorado University
Dept. 1499
200 W. Lake St.
80523 Fort Collins
USA
(970) 491-5573
danr@nrel.colostate.edu

Bruce Wegter
Science center
Hamilton College
198 College Hill Road
13323 Clinton
USA
315 859-4927
bwegter@hamilton.edu

Madhav Machavaram
AWBERC - USEPA
Stable isotope facility
ML 421
26 Martin Luther King W. Drive
45268 Cincinnati
USA
513 - 569 - 7515
Machavaram.Madhav@epa.gov

Jason West
Texas A&M University

TAMU 2138
Dept. of ESSM
77843 College Station
USA
979-845-3772
jbwest@tamu.edu

David Dettman
Environmental Isotope Laboratory
Univ. of Arizona
Dept. of Geosciences
1040 E. Fourth Street
85721 Tucson AZ
USA
+1 (520) 621-4618
dettman@email.arizona.edu

Peter Swart
Marine Geology
Univeristy of Miami
4600 Rickenbacker Causeway
33149 Miami
USA
+1 305 361 4103
pswart@rsmas.miami.edu

Dachun Zhang
IRMS facility, ESS Dept.
University of California
2224 Croul Hall
92697 Irvine

USA
(949) 824-3874
zhangd@uci.edu

James Ehleringer
Department of Biology
University of Utah
257 South 1400 East
84112 Salt Lake City
USA
801-581-7623
jim.ehleringer@utah.edu

William Rugh
Integrated Stable Isotope Research Facility
US EPA, NHEERL
Western Ecology Division
200 SW 35th St.
97333 Corvallis
USA
(541) 754-4604
Rugh.William@epamail.epa.gov

Kien Nguyen Chinh
Department of Isotope Hydrology
Centre for Nuclear Techniques
217 Nguyen Trai, Quan 1
70000 Ho Chi Minh City
Vietnam
0084 8 38356568
nkienchinh@gmail.com

APPENDIX III. WICO2011 ANNOUNCEMENTS

The following announcements were made through IAEA Website www.iaea.org/water, ISOGEOCHEM News Group on Isotope geochemistry and emails to all participants of the former intercomparisons.

Announcement of WICO 2011, IAEA Water Isotope Interlaboratory Comparison

The Isotope Hydrology Laboratory of the International Atomic Energy Agency is organising the fourth interlaboratory comparison for laboratories engaged in stable isotope analyses (^{18}O & ^2H) of water samples. The exercise is part of the IAEA Water Resources Programme to promote quality assurance and to improve the reliability and comparability of stable isotope measurements. Those willing to join the exercise should confirm their participation by sending the complete postal address of the participating institution with name of the investigator, e-mail, fax number and homepage (if available) to the Acting Unit Head of Isotope Hydrology Laboratory with the subject field "WICO 2011": *Manzoor Choudhry, Isotope Hydrology Laboratory, International Atomic Energy Agency, P.O. Box 100, Wagramerstrasse 5, A-1400 Vienna, Austria; fax:43-1-26007; e-mail: isotope.hydrology.lab@iaea.org*. Detailed plan will be provided later.

Progress in WICO 2011, IAEA Water Isotope Interlaboratory Comparison organized by the Isotope Hydrology Laboratory of Water Resources Programme

In response to our announcement, 173 laboratories have been registered to participate in the fourth interlaboratory comparison exercise on stable isotope analyses (^{18}O & ^2H) of water samples. The samples along with covering letter and data reporting sheet have been dispatched to all the laboratories, which are expected to arrive soon. Electronic version of the data reporting sheet has been sent to each laboratory and it can also be downloaded from our website www.iaea.org/water. Please use this data reporting sheet to submit the results for each method used for analysis. The participants are requested to send the report electronically and a hard copy to the address given below before 31st August, 2011.

Manzoor Choudhry
Isotope Hydrology Laboratory,
International Atomic Energy Agency,
P.O. Box 100, Wagramerstrasse 5,
A-1400 Vienna, Austria;
Phone: +43 1 2600 21758
Fax: +43 1 26007;
Email: isotope.hydrology.lab@iaea.org

APPENDIX IV. LABORATORY REPORTING SHEET

**FOURTH IAEA INTERLABORATORY COMPARISON
FOR STABLE ISOTOPE ANALYSES OF WATER**

4th IAEA INTERLABORATORY COMPARISON FOR STABLE ISOTOPE ANALYSES OF WATER	
Laboratory ID	
Laboratory name	
Person in charge	
Reporting date	

Methods:

Analyte	Method	Sample preparation/equilibration	Analyzer
$\delta^{18}\text{O}$	1		
	2		
	3		
$\delta^2\text{H}$	1		
	2		
	3		

Remarks

--

Results

Sample	Method	$\delta^{18}\text{O}$			$\delta^2\text{H}$		
		‰	unc.	n	‰	unc.	n
OH-13	1						
	2						
	3						
OH-14	1						
	2						
	3						
OH-15	1						
	2						
	3						
OH-16	1						
	2						
	3						

Individual results

Sample	Analysis	$\delta^{18}\text{O}$			$\delta^2\text{H}$		
		Method 1	Method 2	Method 3	Method 1	Method 2	Method 3
OH13	1						
	2						
	3						
	4						
	5						
	6						
	7						
	8						
	9						
	10						
OH14	1						
	2						
	3						
	4						
	5						
	6						
	7						
	8						
	9						
	10						
OH15	1						
	2						
	3						
	4						
	5						
	6						
	7						
	8						
	9						
	10						
OH16	1						
	2						
	3						
	4						
	5						
	6						
	7						
	8						
	9						
	10						

Laboratory Manager.....
Date:.....

Please send this reporting sheet as electronic version to:

Manzoor Choudhry
Team Leader, Isotope Hydrology Laboratory
International Atomic Energy Agency
P.O.Box 100 A-1400 Vienna AUSTRIA
Phone: +43-1-2600-21758/21766 Fax: +43-1-2600-7
e-mail (SMTP) : isotope.hydrology.lab@iaea.org

Explanations

Field	Information required	Example
Analyzer	Equilibration system	Brand, custom made
	Mass Spectrometer	Brand, type (dual inlet / continuous flow)
	Laser Spectroscopic Analyzer	Brand, settings (No. of injections)
Detail of sample preparation for each method	Volume of water sample used	4 mL, 10 μ L
	Transfer	direct, capsule
	Reaction (when applicable)	Zinc, 2 hours at 500°C, Glassy Carbon at 1050°C
	Equilibration	4 hours at 25°C
	Separation (when applicable)	GC at 80°C
Normalization	Type of lab standards	How are the lab standards calibrated? Using IAEA VSMOW and SLAP or any other standards
	Number of Standards	1-standard or 2-standard normalisation
	Values of lab. standards	...‰ vs VSMOW normalized on VSMOW-SLAP scale
	Sequence of samples and standards	e.g. calibration standard1, 2, control, sample1,2 etc.
	Calculations method	formulas / templates
	Date of last calibration	Date the lab standards were last calibrated using VSMOW SLAP GISP
Final results	Average of n measurements	...‰ vs VSMOW (2 decimal place for $\delta^{18}\text{O}$ and 1 decimal place for $\delta^2\text{H}$)
Uncertainty	Long term sd of laboratory standards	...‰ (2 decimal place for $\delta^{18}\text{O}$ and 1 decimal place for $\delta^2\text{H}$)
	Standard deviation of individual measurements	...‰ (2 decimal place for $\delta^{18}\text{O}$ and 1 decimal place for $\delta^2\text{H}$)
Individual results	Results of all single measurements.	Ratio ($^{18}\text{O}/^2\text{H}$) vs working standard in case of IRMS
		Please attach machine raw data electronically, eg. LGR: H ₂ O_yyyymmdd.xxx.txt, Picarro: hbdsxx_isowater_yyyymmdd_xxxxx.csv